

State of North Dakota)
)ss
County of Burleigh)

I, Charles E. Eder, do hereby certify that I am a duly-appointed State Toxicologist for the State of North Dakota and an official custodian of the records and files of the office thereof, that I have carefully compared the

Ethanol Breath Standard Analytical Report, Lot No. 08513100A1, Expiration 05/01/2015

hereto attached with the respective original as the same appears of record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, and find the same to be a true and correct copy thereof and of the whole thereof. In witness whereof I have set my hand at the city of Bismarck, in said county this:

20th day of SEPTEMBER, 2013




Charles E. Eder, State Toxicologist

State of North Dakota)
)ss
County of Burleigh)

On this 20 day of September, 2013, before me personally appeared Charles E. Eder, known to me to be a State Toxicologist for the State of North Dakota, and acknowledged to me that he/she has executed the same.

Subscribed to and sworn before me this:

20 day of September, 2013



Stacie L. Fleck-Merkel, Notary Public, State of North Dakota
My Commission Expires August 16, 2017

(SEAL)

ETHANOL BREATH STANDARD ANALYTICAL REPORT

Ethanol Breath Standard Lot Number 08513100A1 Expiration Date 5/1/2015

This standard was analyzed by ILMO Specialty Gases with a reported result of 260.5 ppm which is the equivalent of 0.100 AC of Ethanol in Nitrogen. ILMO Specialty Gases has provided a Certificate of Analysis traceable to N.I.S.T. SRM Ethanol Standards.

A proper result for the standard test using a cylinder of this lot number would be the range of 0.095 to 0.105 g ethanol/210 L of vapor (g/100 ml of blood or g/210 L of end expiratory breath).

The Intoxilyzer® will print out the value of the standard test in 3 digits on Intoxilyzer® Test Record (Form 106-I8000).

The number of cylinders sent to each location will be based on need. The standard may be used until the date of expiration as indicated by the manufacturer's Certificate of Analysis.



Charles E. Eder, State Toxicologist

20 SEP 13

Date Approved



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis


Certificate ID: 4699
Part #: BAC105L100T
Cylinder Size: 105L
Lot Number: 08513100A1
Expiration: 5/1/2015

0.100 BAC (For use with breath alcohol testing instruments)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

| Component: | Concentration: | Accuracy: | Method: |
|------------|----------------|--------------------------|---------|
| Ethanol | 260.5 ppm | +/- 0.002 or 2% | NDIR |
| Nitrogen | Balance | BAC whichever is greater | |

*NIST Standard Reference Material
Cylinder No. CC157787 / Job No. 13029
Certified 218.1 µmol/mol Ethanol in Nitrogen
for ILMO Products Co., Jacksonville, IL


Specialty Gas Lab Tech

04/08/13
Date

Distributed by: CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

