


State of North Dakota)
)ss
County of Burleigh)

I, Charles E. Eder, do hereby certify that I am the duly-appointed State Toxicologist for the State of North Dakota and an official custodian of the records and files of the office thereof, that I have carefully compared the

Intoximeters Alco-Sensor FST® Ethanol Breath Standard Analytical Report, Lot No. AG630601, Expiration 11/01/18 (3/27/17)

hereto attached with the respective original as the same appears of record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, and find the same to be a true and correct copy thereof and of the whole thereof. In witness whereof I have set my hand at the city of Bismarck, in said county this:

27th day of MARCH, 2017




Charles E. Eder, State Toxicologist

State of North Dakota)
)ss
County of Burleigh)


On this 27th day of March, 2017, before me personally appeared Charles E. Eder, known to me to be a State Toxicologist for the State of North Dakota, and acknowledged to me that he has executed the same.

Subscribed to and sworn before me this:

27th day of March, 2017



Jeremy Spaeth, Notary Public, State of North Dakota
My Commission Expires November 05, 2019



(SEAL)

**INTOXIMETERS ALCO-SENSOR FST®
ETHANOL BREATH STANDARD ANALYTICAL REPORT**

Ethanol Breath Standard Lot Number AG630601 Expiration Date 11/01/2018

This standard was analyzed by AIRGAS USA LLC (LAB) with a reported result of 223 ppm which is the equivalent of 0.082 BrAC of Ethanol in Nitrogen. AIRGAS USA LLC (LAB) has provided a Certificate of Analysis traceable to N.I.S.T. RGM Ethanol Standards.

The number of cylinders sent to each location will be based on need. The standard may be used until the date of expiration as indicated by the manufacturer's Certificate of Analysis.



Charles E. Eder, State Toxicologist

27 MAR 2017

Date Approved



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 1-Nov-2016

Lot # AG630601 Model 108caccd

Exp. Date

1-Nov-2018

Cyl. Type

108

Component

Ethanol

Nitrogen

Certified Concentration

0.082 ± 0.002 BrAC (223 ppm)

Balance

Certification Traceable to N.I.S.T. RGM Ethanol Standards:

Serial No.

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

Concentration

391.8 ppm

259.8 ppm

209.0 ppm

103.7 ppm

52.22 ppm

Serial No.

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

Concentration

392.5 ppm

258.9 ppm

208.9 ppm

104.9 ppm

52.94 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2016.11.04 09:24:44 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 2989.01