

State of North Dakota)
)ss
County of Burleigh)

I, Charles E. Eder, do hereby certify that I am the duly-appointed State Toxicologist for the State of North Dakota and an official custodian of the records and files of the office thereof, that I have carefully compared the

Intoximeters Alco-Sensor FST® Ethanol Breath Standard Analytical Report, Lot No. AG009001, Expiration 03/30/22 (06/09/20)

hereto attached with the respective original as the same appears of record on file in the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, and find the same to be a true and correct copy thereof and of the whole thereof. In witness whereof I have set my hand at the city of Bismarck, in said county this:

9th day of JUNE, 2020




Charles E. Eder, State Toxicologist

State of North Dakota)
)ss
County of Burleigh)


On this 9th day of June, 2020, before me personally appeared Charles E. Eder, known to me to be the State Toxicologist for the State of North Dakota, and acknowledged to me that he has executed the same.

Subscribed to and sworn before me this:

9th day of June, 2020



Deanna Dailey, Notary Public, State of North Dakota
My Commission Expires March 23, 2023



(SEAL)

**INTOXIMETERS ALCO-SENSOR FST®
ETHANOL BREATH STANDARD ANALYTICAL REPORT**

Ethanol Breath Standard Lot Number AG009001 Expiration Date 03/30/2022

This standard was analyzed by AIRGAS USA LLC (LAB) with a reported result of 223 ppm which is the equivalent of 0.082 BrAC of Ethanol in Nitrogen. AIRGAS USA LLC (LAB) has provided a Certificate of Analysis traceable to N.I.S.T. RGM Ethanol Standards.

The number of cylinders sent to each location will be based on need. The standard may be used until the date of expiration as indicated by the manufacturer's Certificate of Analysis.



Charles E. Eder, State Toxicologist

6.9.2020
Date Approved



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 30-Mar-2020

Lot # AG009001 Model 108cacd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
30-Mar-2022	108	Ethanol Nitrogen	0.082 ± 0.002 BrAC (223 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2020.03.30 19:43:14 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07