

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004188

Location: TOXL

- |  | Check When Done |
|--|-----------------|
| A. Is the warm-up time less than 20 minutes?   | <u>✓</u>        |
| B. Is a three-pronged grounded outlet used?  | <u>✓</u>        |
| C. Is the breath tube heated?  | <u>✓</u>        |
| D. Is the diagnostic check complete?   | <u>✓</u>        |
| E. Is the time, date, and year correct (re-set if necessary)?  | <u>✓</u>        |
| F. Print test. (Attach test record.)   | <u>✓</u>        |
| G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5132</u> Lot No. <u>    </u> Rep AC <u>0.021</u>   | <u>✓</u>        |
| H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)<br>Sim SN <u>DR5134</u> Lot No. <u>    </u> Rep AC <u>0.294</u>  | <u>✓</u>        |
| I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.<br>(Attach test record.)<br>Sim SN <u>DR3846</u> Lot No. <u>    </u> Rep AC <u>0.10 AC + acetone</u> | <u>✓</u>        |
| J. RFI check. Run CMS mode. Key radio on first room air.<br>(Attach test record.)  | <u>✓</u>        |
| K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.<br>(Attach test record.)  |                 |

Lot No. <u>15410080A1</u>	Cyl No. <u>12</u>	Exp. Date <u>07/01/2012</u>
Test 1 <u>0.080</u>	Test 4 <u>0.080</u>	Test 7 <u>0.080</u>
Test 2 <u>0.079</u>	Test 5 <u>0.079</u>	Test 8 <u>0.079</u>
Test 3 <u>0.079</u>	Test 6 <u>0.079</u>	Test 9 <u>0.080</u>
Average <u>0.079</u>		

L. Tank pressure, Level 3 Func D Sub G P: 405 psi Reg 400 psi

M. Remarks/Maintenance record: Annual inspection, replaced gas cyl regulator.

01 Apr 2011  
Date

Dob Kashur  
Field Inspector's Signature

1 Apr 2011  
Date

Murray Blum  
Reviewed by

80-004188  
P. 118-1126  
FM



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004188  
Location = TOXL                8164.13.00 06/09  
04/01/2011                        15:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:24
02 Std. Sol.	0.020	15:25
03 Room Air	0.000	15:25
04 Std. Sol.	0.020	15:26
05 Room Air	0.000	15:27
06 Std. Sol.	0.020	15:27
07 Room Air	0.000	15:28

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

low Ae 0.021AC

Form 106-I8000

80-004188  
p. 120 of 126  
FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004188  
Location = TOXL      8164.13.00 06/09  
04/01/2011      15:14

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:15
02 Std. Sol.	0.286	15:16
03 Room Air	0.000	15:16
04 Std. Sol.	0.288	15:17
05 Room Air	0.000	15:18
06 Std. Sol.	0.288	15:18
07 Room Air	0.000	15:19

08 Sim Temp = 34.0°C

Simul Ser No = 0

Std Sol No = 0

County = 08

Oper No. = 77777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

Linearity check 0.294 AC  
~~Interferent check 0.294 AC~~

Form 106-I8000

80-004188  
P121 of 126  
FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004188  
Location = TOXL                8164.13.00 06/09  
04/01/2011                        15:07

Test	AC	Time
01 Room Air	0.000	15:08
02 *Subject Test	INT*	15:08
03 Room Air	0.000	15:09

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 15410080A1  
Cyl No = 12  
Expiration Date = 07/01/2012  
County = 08                          Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*Interferent check 0.10 Ac<sup>OK</sup> + acetone*

Form 106-I8000

80-004188  
P.122 of 126  
VM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004188  
Location = TOXL                    8164.13.00 06/09  
04/01/2011                            15:09

Test	AC	Time
01 Diagnostic	OK	15:10
02 Room Air	RFI*	15:10
03 Room Air	0.000	15:11

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 15410080A1  
Cyl No = 12  
Expiration Date = 07/01/2012  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
DEB KASHUR

Remarks:

*RFI test*

Form 106-I8000

80-004188  
p. 123 of 126  
rjm

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004188  
Location = TOXL      8164.13.00 06/09  
04/01/2011      14:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:46
02 Std. Gas	0.080	14:46
03 Room Air	0.000	14:47
04 Std. Gas	0.079	14:47
05 Room Air	0.000	14:48
06 Std. Gas	0.079	14:48
07 Room Air	0.000	14:48

Lot No = 15410080A1  
Cyl No = 12  
Exp Date = 07/01/2012  
County = 08      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

80-004188  
P124 of 126  
FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004188  
Location = TOXL      8164.13.00 06/09  
04/01/2011      14:57

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:58
02 Std. Gas	0.080	14:58
03 Room Air	0.000	14:59
04 Std. Gas	0.079	14:59
05 Room Air	0.000	15:00
06 Std. Gas	0.079	15:00
07 Room Air	0.000	15:01

Lot No = 15410080A1  
Cyl No = 12  
Exp Date = 07/01/2012  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

80-004188  
p. 125 of 126  
FM



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004188  
Location = TOXL      8164.13.00 06/09  
04/01/2011      15:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:04
02 Std. Gas	0.080	15:04
03 Room Air	0.000	15:04
04 Std. Gas	0.079	15:05
05 Room Air	0.000	15:05
06 Std. Gas	0.080	15:06
07 Room Air	0.000	15:06

Lot No = 15410080A1  
Cyl No = 12  
Exp Date = 07/01/2012  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004188  
P. 126 of 126  
YTH