

INTOXILYZER 8000 INSPECTION

Serial No.: 80-004951

Location: Newtown

- Check When Done
- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5132 Lot No. Rep AC 0.021 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5115 Lot No. Rep AC 0.298 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR5186 Lot No. Rep AC 0.100 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. <u>709352</u>	Cyl No. <u>36</u>	Exp. Date <u>12/23/11</u>
Test 1 <u>0.078</u>	Test 4 <u>0.078</u>	Test 7 <u>0.078</u>
Test 2 <u>0.077</u>	Test 5 <u>0.078</u>	Test 8 <u>0.078</u>
Test 3 <u>0.078</u>	Test 6 <u>0.077</u>	Test 9 <u>0.078</u>
Average <u>0.078</u>		

L. Tank pressure, Level 3 Func D Sub G P: 821 psi Reg 838 psi

M. Remarks/Maintenance record: Annual inspection, changed date in 2E

05 May 2010
Date

Deb Kashur
Field Inspector's Signature

15 May 2010
Date

Mary Penn
Reviewed by

80-004951
P. 38/46 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004951
Location = NEWT 8164.13.00 06/09
~~05/15/2010~~ 08:49
05/

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 709352
Standard Cyl #? 36
Standard Expiration? 12/23/2011
Oper No? 777777

Flow Cal. Date: 01/19/2010
Slope 698
Intercept -658617

IR Calibration Date: 01/19/2010
 3um 9um

0th Coef(*100):	-24390	-26496
1st Coef(*100):	258099	132326
2nd Coef(*100):	1745	1171
H2O adj (mg/l*10k):	509	459

***** Printer Test End *****

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Print test

Form 106-I8000

80-004951
P. 39/46 PM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004951
Location = NEWT 8164.13.00 06/09
05/05/2010 09:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:07
02 Std. Sol.	0.017	09:08
03 Room Air	0.000	09:09
04 Std. Sol.	0.017	09:10
05 Room Air	0.000	09:10
06 Std. Sol.	0.018	09:11
07 Room Air	0.000	09:11

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 60

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Low AC 0.021 AC

Form 106-I8000

80-004951
P. 40/46 JAL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004951
Location = NEWT 8164.13.00 06/09
05/05/2010 09:12

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:13
02 Std. Sol.	0.302	09:14
03 Room Air	0.000	09:14
04 Std. Sol.	0.302	09:15
05 Room Air	0.000	09:16
06 Std. Sol.	0.301	09:16
07 Room Air	0.000	09:17

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 60

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.298 AC

Form 106-18000

80-004951
P. 41/46 *AK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004951
Location = NEWT 8164.13.00 06/09
05/05/2010 09:05

Test	AC	Time
01 Room Air	0.000	09:06
02 *Subject Test	INT*	09:06
03 Room Air	0.000	09:06

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 709352
Cyl No = 36
Expiration Date = 12/23/2011
County = 60 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check

Form 106-18000

*0.100 AC
+ acetone*

*80-004951
P. 42/46 FM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004951
Location = NEWT 8164.13.00 06/09
05/05/2010 09:03

Test	AC	Time
01 Diagnostic	OK	09:04
02 Room Air	RFI*	09:04
03 Room Air	0.000	09:05

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 709352
Cyl No = 36
Expiration Date = 12/23/2011
County = 60 Oper No. = 777777

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

RFI check

Form 106-I8000

80-004951
p. 43/46 *rm*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004951
Location = NEWT 8164.13.00 06/09
05/05/2010 08:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:53
02 Std. Gas	0.078	08:54
03 Room Air	0.000	08:54
04 Std. Gas	0.077	08:55
05 Room Air	0.000	08:55
06 Std. Gas	0.078	08:55
07 Room Air	0.000	08:56

Lot No = 709352

Cyl No = 36

Exp Date = 12/23/2011

County = 60

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004951
P. 44/46. *rm*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004951
Location = NEWT 8164.13.00 06/09
05/05/2010 08:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:57
02 Std. Gas	0.078	08:57
03 Room Air	0.000	08:58
04 Std. Gas	0.078	08:58
05 Room Air	0.000	08:59
06 Std. Gas	0.077	08:59
07 Room Air	0.000	09:00

Lot No = 709352
Cyl No = 36
Exp Date = 12/23/2011
County = 60 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000

80-004951
P. 45/46 PM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004951
Location = NEWT 8164.13.00 06/09
05/05/2010 09:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:00
02 Std. Gas	0.078	09:01
03 Room Air	0.000	09:01
04 Std. Gas	0.078	09:02
05 Room Air	0.000	09:02
06 Std. Gas	0.078	09:02
07 Room Air	0.000	09:03

Lot No = 709352
Cyl No = 36
Exp Date = 12/23/2011
County = 60 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

calibration check 0.080 AC

Form 106-I8000

80-004951
P. 46/46 RM