

INTOXILYZER® 8000 INSPECTION

(BIA
FORT
BERTHOLD)

Serial No.: 80-005945

Location: NEWT

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5191 Lot No. 15160 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 16060 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR5190 Lot No. ICS4 Solution 0.10AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 29614080A2 Cyl No. 029 Exp. Date 12/05/2016
Test 1 0.079 Test 4 0.079 Test 7 0.080
Test 2 0.080 Test 5 0.079 Test 8 0.079
Test 3 0.079 Test 6 0.079 Test 9 0.079
Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 798 psi Reg 800 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/3/2016
Date

Charles E. Edin
Field Inspector's Signature

05 May 16
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
 NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
 North Dakota Model 8000 SN 80-005945
 Location = NEWT 8164.13.00 06/09
 05/03/2016 14:11

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-==|
 ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-==|
 ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup
 Data Entry Mode: Enabled
 Start Test Sequence: DABACABA
 Display Prelim Rslt? Yes
 Display Third Digit? Yes
 Inhib Printer(Y/N)? No
 Display Volume? No
 Disable On Memfull? Yes
 # of Print Copies? 1
 Select Std (D/W/I)? Dry
 Standard Value? 0.080
 Standard Lot #? 29614080A2
 Standard Cyl #? 29
 Standard Expiration? 12/05/2016
 Oper No? 666666

Flow Cal. Date: 06/20/2013
 Slope 637
 Intercept -566407

IR Calibration Date: 05/22/2015

	3um	9um
0th Coef(*100):	-21213	-23093
1st Coef(*100):	279171	139292
2nd Coef(*100):	2898	1291
H2O adj(mg/l*10k):	332	319

***** Printer Test End *****



Operator Signature
 CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.13.00 06/09
05/03/2016 14:17

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:18
02 Std. Sol.	0.018	14:19
03 Room Air	0.000	14:19
04 Std. Sol.	0.019	14:20
05 Room Air	0.000	14:21
06 Std. Sol.	0.019	14:21
07 Room Air	0.000	14:22

08 Sim Temp = 34.0°C

Simul Ser No = DR5191
Std Sol No = 15160
County = 60 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.13.00 06/09
05/03/2016 14:29

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:29
02 Std. Sol.	0.310	14:30
03 Room Air	0.000	14:31
04 Std. Sol.	0.312	14:31
05 Room Air	0.000	14:32
06 Std. Sol.	0.313	14:33
07 Room Air	0.000	14:33

08 Sim Temp = 34.0°C

Simul Ser No = DR3847

Std Sol No = 16060

County = 60

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.13.00 06/09
05/03/2016 14:40

Test	AC	Time
01 Room Air	0.000	14:41
02 *Subject Test	INT*	14:41
03 Room Air	0.000	14:42

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 029
Expiration Date = 12/05/2016
County = 60 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.13.00 06/09
05/03/2016 14:42

Test	AC	Time
01 Diagnostic	OK	14:43
02 Room Air	0.000	14:43
03 *Subject Test	RFI*	14:44
04 Room Air	0.000	14:44

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 29614080A2
Cyl No = 029
Expiration Date = 12/05/2016
County = 60 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.13.00 06/09
05/03/2016 14:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:37
02 Std. Gas	0.079	14:37
03 Room Air	0.000	14:38
04 Std. Gas	0.080	14:38
05 Room Air	0.000	14:38
06 Std. Gas	0.079	14:39
07 Room Air	0.000	14:39

Lot No = 29614080A2
Cyl No = 029
Exp Date = 12/05/2016
County = 60 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.13.00 06/09
05/03/2016 14:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:45
02 Std. Gas	0.079	14:45
03 Room Air	0.000	14:46
04 Std. Gas	0.079	14:46
05 Room Air	0.000	14:47
06 Std. Gas	0.079	14:47
07 Room Air	0.000	14:48

Lot No = 29614080A2
Cyl No = 029
Exp Date = 12/05/2016
County = 60 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.13.00 06/09
05/03/2016 14:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:48
02 Std. Gas	0.080	14:49
03 Room Air	0.000	14:49
04 Std. Gas	0.079	14:49
05 Room Air	0.000	14:50
06 Std. Gas	0.079	14:50
07 Room Air	0.000	14:51

Lot No = 29614080A2
Cyl No = 029
Exp Date = 12/05/2016
County = 60 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000