

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-004184 Location: CROS (DIVIDE COUNTY SO.)

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN PR5115 Lot No. 12111 Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN PR5143 Lot No. 13080 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3375 Lot No. ICS#2 Rep AC 0.100 AC + 0.05% ACETONE ✓
 - H. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 16512090A2 Cyl No. 051 Exp. Date 08/01/2014
Test 1 0.078 Test 4 0.079 Test 7 0.078
Test 2 0.079 Test 5 0.079 Test 8 0.078
Test 3 0.079 Test 6 0.078 Test 9 0.078
Average 0.078

J. Tank pressure, Level 3 Func D Sub G P: 887 psi Reg 900 psi

K. Remarks/Maintenance record: REPLACED BREATH HOSE AND SIMULATOR RETURN
O-RING, ANNUAL INSPECTION

28 MAY 14
Date

29 May 14 ✓
Date

Charles E. Eden
Field Inspector's Signature

Deb Kashner
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = CROS 8164.13.00 06/09
05/28/2014 13:00

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 51
Standard Expiration? 08/01/2014
Oper No? 666666

Flow Cal. Date: 07/30/2010
Slope 804
Intercept -925008

IR Calibration Date: 05/08/2013
 3um 9um

0th Coef(*100): -15562 -18504
1st Coef(*100): 268750 136542
2nd Coef(*100): 2481 1345
H2O adj(mg/l*10k): 568 556

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = CROS 8164.13.00 06/09
05/28/2014 13:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:21
02 Std. Sol.	0.020	13:21
03 Room Air	0.000	13:22
04 Std. Sol.	0.020	13:23
05 Room Air	0.000	13:23
06 Std. Sol.	0.020	13:24
07 Room Air	0.000	13:24

08 Sim Temp = 34.0°C

Simul Ser No = DR5115
Std Sol No = 12111
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = CROS 8164.13.00 06/09
05/28/2014 13:26

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:27
02 Std. Sol.	0.312	13:28
03 Room Air	0.000	13:28
04 Std. Sol.	0.314	13:29
05 Room Air	0.000	13:30
06 Std. Sol.	0.314	13:30
07 Room Air	0.000	13:31

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 12

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = CROS 8164.13.00 06/09
05/28/2014 13:03

Test	AC	Time
01 Room Air	0.000	13:04
02 *Subject Test	INT*	13:05
03 Room Air	0.000	13:06

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 51
Expiration Date = 08/01/2014
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

*INTERFERENT
CHECK
0.100 AC + 0.05% ACETONE*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = CROS 8164.13.00 06/09
05/28/2014 13:00

Test	AC	Time
01 Diagnostic	OK	13:02
02 Room Air	RFI*	13:02
03 Room Air	0.000	13:02

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 16512080A2
Cyl No = 51
Expiration Date = 08/01/2014
County = 12 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = CROS 8164.13.00 06/09
05/28/2014 13:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:07
02 Std. Gas	0.078	13:08
03 Room Air	0.000	13:08
04 Std. Gas	0.079	13:09
05 Room Air	0.000	13:09
06 Std. Gas	0.079	13:09
07 Room Air	0.000	13:10

Lot No = 16512080A2
Cyl No = 51
Exp Date = 08/01/2014
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = CROS 8164.13.00 06/09
05/28/2014 13:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:11
02 Std. Gas	0.079	13:11
03 Room Air	0.000	13:12
04 Std. Gas	0.079	13:12
05 Room Air	0.000	13:13
06 Std. Gas	0.078	13:13
07 Room Air	0.000	13:13

Lot No = 16512080A2

Cyl No = 51

Exp Date = 08/01/2014

County = 12

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CAIBRATION
CHECK
0.0804C

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = CROS 8164.13.00 06/09
05/28/2014 13:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:35
02 Std. Gas	0.078	13:35
03 Room Air	0.000	13:35
04 Std. Gas	0.078	13:36
05 Room Air	0.000	13:36
06 Std. Gas	0.078	13:37
07 Room Air	0.000	13:37

Lot No = 16512080A2
Cyl No = 51
Exp Date = 08/01/2014
County = 12 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION
CHECK
0.080 AC

Form 106-I8000