

INTOXILYZER 8000 INSPECTION

Serial No.: 80-003069

Location: ROLL (ROULETTE C.S.O.)

- Check When Done
- A. Is the warm-up time less than 20 minutes? ✓
 - B. Is a three-pronged grounded outlet used? ✓
 - C. Is the breath tube heated? ✓
 - D. Is the diagnostic check complete? ✓
 - E. Is the time, date, and year correct (re-set if necessary)? ✓
 - F. Print test. (Attach test record.) ✓
 - G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR3455 Lot No. 12010 Rep AC 0.020 ✓
 - H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3847 Lot No. 11110 Rep AC 0.300 ✓
 - I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3377 Lot No. — Rep AC 0.100 AC ✓
 - J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
 - K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 05911 080A1 Cyl No. 28 Exp. Date 2/1/2013
Test 1 0.077 Test 4 0.076 Test 7 0.077
Test 2 0.077 Test 5 0.077 Test 8 0.077
Test 3 0.077 Test 6 0.076 Test 9 0.078
Average 0.077

L. Tank pressure, Level 3 Func D Sub G P: 900 psi Reg 750 psi

M. Remarks/Maintenance record: ANNUAL INSPECTION

14 JUN 12
Date

Charles Elden
Field Inspector's Signature

15 June 12
Date

Deb Kashur
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = ROLL 8164.13.00 06/09
06/14/2012 14:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:59
02 Std. Sol.	0.018	15:00
03 Room Air	0.000	15:01
04 Std. Sol.	0.018	15:01
05 Room Air	0.000	15:02
06 Std. Sol.	0.018	15:03
07 Room Air	0.000	15:03

08 Sim Temp = 34.0°C

Simul Ser No = 000000
Std Sol No = 0000
County = 40 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Low AC*
0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = ROLL 8164.13.00 06/09
06/14/2012 15:09

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:09
02 Std. Sol.	0.291	15:10
03 Room Air	0.000	15:11
04 Std. Sol.	0.290	15:11
05 Room Air	0.000	15:12
06 Std. Sol.	0.290	15:13
07 Room Air	0.000	15:13

08 Sim Temp = 34.0°C

Simul Ser No = 000000
Std Sol No = 0000
County = 40 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *LINEARITY TEST*
0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = ROLL 8164.13.00 06/09
06/14/2012 14:49

Test	AC	Time
01 Room Air	0.000	14:50
02 *Subject Test	INT*	14:51
03 Room Air	0.000	14:51

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = MNBVCXZLKJHG
Dr. Lic. = ND/DIS821456
Lot No = 05911080A1
Cyl No = 28
Expiration Date = 02/01/2013
County = 40 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *INTERFERENT CHECK*
0.100AC + ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = ROLL 8164.13.00 06/09
06/14/2012 14:46

Test	AC	Time
01 Diagnostic	OK	14:47
02 Room Air	RFI*	14:47
03 Room Air	0.000	14:48

*Invalid Test
 Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = QWERTYUIOPASDFG
Dr. Lic. = ND/DIS821456
Lot No = 05911080A1
Cyl No = 28
Expiration Date = 02/01/2013
County = 40 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: *RFI CHECK*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = ROLL 8164.13.00 06/09
06/14/2012 15:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:16
02 Std. Gas	0.077	15:16
03 Room Air	0.000	15:17
04 Std. Gas	0.077	15:17
05 Room Air	0.000	15:18
06 Std. Gas	0.077	15:18
07 Room Air	0.000	15:18

Lot No = 05911080A1
Cyl No = 28
Exp Date = 02/01/2013
County = 40 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = ROLL 8164.13.00 06/09
06/14/2012 15:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:19
02 Std. Gas	0.076	15:20
03 Room Air	0.000	15:20
04 Std. Gas	0.077	15:20
05 Room Air	0.000	15:21
06 Std. Gas	0.076	15:21
07 Room Air	0.000	15:22

Lot No = 05911080A1
Cyl No = 28
Exp Date = 02/01/2013
County = 40 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.080 AL


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = ROLL 8164.13.00 06/09
06/14/2012 15:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:23
02 Std. Gas	0.077	15:23
03 Room Air	0.000	15:23
04 Std. Gas	0.077	15:24
05 Room Air	0.000	15:24
06 Std. Gas	0.078	15:25
07 Room Air	0.000	15:25

Lot No = 05911080A1
Cyl No = 28
Exp Date = 02/01/2013
County = 40 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC