

INTOXILYZER 8000 INSPECTION

Serial No.: 80-003069

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR3803 Lot No. 13070 Rep AC 0.020 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5143 Lot No. 13080 Rep AC 0.300 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR3846 Lot No. 1CS2 Rep AC 0.10 AC + acetone (0.05%) ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 23411080 A1 Cyl No. 38 Exp. Date 10/01/2013

Test 1 0.079 Test 4 0.079 Test 7 0.079

Test 2 0.080 Test 5 0.080 Test 8 0.079

Test 3 0.079 Test 6 0.080 Test 9 0.080

Average 0.079

L. Tank pressure, Level 3 Func D Sub G P: 277 psi Reg 270 psi

M. Remarks/Maintenance record: Annual inspection

26 Sept 2013
Date

Deb Kashur
Field Inspector's Signature

26 SEP 13
Date

Charles E. Ed
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.13.00 06/09
09/26/2013 09:42

***** Printer Test *****

abcdefghijklmnopqrstu vwxyz1234567890-=_|
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstu vwxyz1234567890-=_|
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 23411080A1
Standard Cyl #? 38
Standard Expiration? 10/01/2013
Oper No? 777777

Flow Cal. Date: 09/17/2013
Slope 678
Intercept -734495

IR Calibration Date: 09/17/2013
 3um 9um

0th Coef(*100): -18028 -10092
1st Coef(*100): 282280 137019
2nd Coef(*100): 2090 1032
H2O adj(mg/l*10k): 503 189

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.13.00 06/09
09/26/2013 10:00

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:01
02 Std. Sol.	0.020	10:02
03 Room Air	0.000	10:02
04 Std. Sol.	0.020	10:03
05 Room Air	0.000	10:03
06 Std. Sol.	0.020	10:04
07 Room Air	0.000	10:05

08 Sim Temp = 34.0°C

Simul Ser No = DR3803

Std Sol No = 13070

County = 08

Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

low AC 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.13.00 06/09
09/26/2013 10:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:07
02 Std. Sol.	0.297	10:08
03 Room Air	0.000	10:08
04 Std. Sol.	0.299	10:09
05 Room Air	0.000	10:09
06 Std. Sol.	0.298	10:10
07 Room Air	0.000	10:11

08 Sim Temp = 34.0°C

Simul Ser No = DR5143

Std Sol No = 13080

County = 08

Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.13.00 06/09
09/26/2013 10:14

Test	AC	Time
01 Room Air	0.000	10:15
02 *Subject Test	INT*	10:15
03 Room Air	0.000	10:16

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Simul Ser No = DR5143
Std Sol No = 13080
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Interferent check 0.10AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.13.00 06/09
09/26/2013 10:17

Test	AC	Time
01 Diagnostic	OK	10:18
02 Room Air	RFI*	10:18
03 Room Air	0.000	10:18

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Simul Ser No = DR5143
Std Sol No = 13080
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.13.00 06/09
09/26/2013 10:19

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:20
02 Std. Gas	0.079	10:20
03 Room Air	0.000	10:21
04 Std. Gas	0.080	10:21
05 Room Air	0.000	10:22
06 Std. Gas	0.079	10:22
07 Room Air	0.000	10:22

Lot No = 23411080A1
Cyl No = 38
Exp Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.13.00 06/09
09/26/2013 10:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:23
02 Std. Gas	0.079	10:24
03 Room Air	0.000	10:24
04 Std. Gas	0.080	10:25
05 Room Air	0.000	10:25
06 Std. Gas	0.080	10:25
07 Room Air	0.000	10:26

Lot No = 23411080A1
Cyl No = 38
Exp Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.13.00 06/09
09/26/2013 10:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:30
02 Std. Gas	0.079	10:31
03 Room Air	0.000	10:31
04 Std. Gas	0.079	10:32
05 Room Air	0.000	10:32
06 Std. Gas	0.080	10:32
07 Room Air	0.000	10:33

Lot No = 23411080A1
Cyl No = 38
Exp Date = 10/01/2013
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check 0.080 AC

Form 106-I8000