

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-005945

Location: NEWT

(NEW TOWN P.D.
BIA FT. BERTHOLD
T.A.T.)

Check When Done

- A. Is the breath tube heated? ✓
- B. Is the diagnostic check complete? ✓
- C. Is the time, date, and year correct (re-set if necessary)? ✓
- D. Print test. (Attach test record.) ✓
- E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5113 Lot No. 201507P Rep AC 0.020 ✓
- F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR3378 Lot No. 16060 Rep AC 0.300 ✓
- G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR3803 Lot No. ICS4 Solution 0.10AC + 0.05% ACETONE ✓
- H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓

I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 17316080A^{CEE}2 Cyl No. 35 Exp. Date 8/5/2018
Test 1 0.080 Test 4 0.079 Test 7 0.079
Test 2 0.080 Test 5 0.079 Test 8 0.079
Test 3 0.079 Test 6 0.079 Test 9 0.079
Average 0.079

J. Tank pressure, Level 3 Func D Sub G P: 904 psi Reg 900 psi

K. Remarks/Maintenance record: ANNUAL INSPECTION

5/16/2017
Date

Charles E. Ed
Field Inspector's Signature

22 May 17
Date

Deb Shanaver
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.14.00 09/16
05/16/2017 09:31

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 17316080A2
Standard Cyl #? 35
Standard Expiration? 08/05/2018
Oper No? 666666

Flow Cal. Date: 06/20/2013
Slope 637
Intercept -566407

IR Calibration Date: 05/22/2015
3um 9um

0th Coef(*100): -21213 -23093
1st Coef(*100): 279171 139292
2nd Coef(*100): 2898 1291
H2O adj(mg/l*10k): 332 319

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.14.00 09/16
05/16/2017 10:07

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:08
02 Std. Sol.	0.016	10:09
03 Room Air	0.000	10:09
04 Std. Sol.	0.017	10:10
05 Room Air	0.000	10:11
06 Std. Sol.	0.017	10:11
07 Room Air	0.000	10:12

08 Sim Temp = 34.0°C

Simul Ser No = DR5113
Std Sol No = 201507F
County = 60 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: Low AC
 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.14.00 09/16
05/16/2017 10:13

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:13
02 Std. Sol.	0.308	10:14
03 Room Air	0.000	10:15
04 Std. Sol.	0.310	10:16
05 Room Air	0.000	10:16
06 Std. Sol.	0.310	10:17
07 Room Air	0.000	10:17

08 Sim Temp = 34.0°C

Simul Ser No = DR3378
Std Sol No = 16060
County = 60 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: HIGH AC
0.300 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.14.00 09/16
05/16/2017 09:46

Test	AC	Time
01 Room Air	0.000	09:47
02 *Subject Test	INT*	09:48
03 Room Air	0.000	09:48

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 17316080A2
Cyl No = 35
Expiration Date = 08/05/2018
County = 60 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *Interferent*
0.10AC + 0.05% ACETONE

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.14.00 09/16
05/16/2017 09:44

Test	AC	Time
01 Diagnostic	OK	09:45
02 Room Air	0.000	09:45
03 *Subject Test	RFI*	09:45
04 Room Air	0.000	09:46

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 17316080A2
Cyl No = 35
Expiration Date = 08/05/2018
County = 60 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: RFI CHECK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.14.00 09/16
05/16/2017 09:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:41
02 Std. Gas	0.080	09:41
03 Room Air	0.000	09:42
04 Std. Gas	0.080	09:42
05 Room Air	0.000	09:43
06 Std. Gas	0.079	09:43
07 Room Air	0.000	09:43

Lot No = 17316080A2
Cyl No = 35
Exp Date = 08/05/2018
County = 60 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.14.00 09/16
05/16/2017 09:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:49
02 Std. Gas	0.079	09:50
03 Room Air	0.000	09:50
04 Std. Gas	0.079	09:50
05 Room Air	0.000	09:51
06 Std. Gas	0.079	09:51
07 Room Air	0.000	09:52

Lot No = 17316080A2
Cyl No = 35
Exp Date = 08/05/2018
County = 60 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005945
Location = NEWT 8164.14.00 09/16
05/16/2017 10:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:19
02 Std. Gas	0.079	10:19
03 Room Air	0.000	10:19
04 Std. Gas	0.079	10:20
05 Room Air	0.000	10:20
06 Std. Gas	0.079	10:21
07 Room Air	0.000	10:21

Lot No = 17316080A2
Cyl No = 35
Exp Date = 08/05/2018
County = 60 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC