

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006491

Location: Pembina Co SO

- Check When Done
- A. Is the breath tube heated? ✓
 - B. Is the diagnostic check complete? ✓
 - C. Is the time, date, and year correct (re-set if necessary)? ✓
 - D. Print test. (Attach test record.) ✓
 - E. Low AC. Use ≤ 0.03 AC in ACA mode. (Attach test record.)
Sim SN DR 5113 Lot No. 261507F Rep AC 0.020 ✓
 - F. High AC. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR 3318 Lot No. 16660 Rep AC 0.300 ✓
 - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.
(Attach test record.)
Sim SN DR 3803 Lot No. 1C55 Solution 0.10AC + 0.05% acetone ✓
 - H. RFI check. Run CMS mode. Key radio during CMS test.
(Attach test record.) ✓
 - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.) 0.080 AC

Lot No. 01316080A2 Cyl No. 28 Exp. Date 02/05/2018 ²⁰¹⁸
Test 1 0.080 Test 4 0.079 Test 7 0.080
Test 2 0.079 Test 5 0.080 Test 8 0.079
Test 3 0.080 Test 6 0.080 Test 9 0.079
Average 0.080

J. Tank pressure, Level 3 Func D Sub G P: 969 psi Reg 980 psi

K. Remarks/Maintenance record: Annual inspection, replaced simulator return o-ring

25 May 17
Date

30 MAY 2017
Date

Deb Shanaver
Field Inspector's Signature

Charles E. Ed
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = CAVA 8164.14.00 09/16
05/25/2017 11:10

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 01316080A2
Standard Cyl #? 28
Standard Expiration? 02/05/2018
Oper No? 777777

Flow Cal. Date: 08/18/2015
Slope 669
Intercept -662544

IR Calibration Date: 08/18/2015
 3um 9um

	3um	9um
0th Coef(*100):	-27273	-22751
1st Coef(*100):	272579	140235
2nd Coef(*100):	1982	1330
H2O adj(mg/l*10k):	664	431

***** Printer Test End *****

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = CAVA 8164.14.00 09/16
05/25/2017 11:15

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:15
02 Std. Sol.	0.017	11:16
03 Room Air	0.000	11:17
04 Std. Sol.	0.017	11:17
05 Room Air	0.000	11:18
06 Std. Sol.	0.018	11:19
07 Room Air	0.000	11:19

08 Sim Temp = 34.0°C

Simul Ser No = DR5113

Std Sol No = 201507F

County = 34

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

low AC 0.020 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = CAVA 8164.14.00 09/16
05/25/2017 11:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:20
02 Std. Sol.	0.302	11:21
03 Room Air	0.000	11:22
04 Std. Sol.	0.302	11:23
05 Room Air	0.000	11:23
06 Std. Sol.	0.303	11:24
07 Room Air	0.000	11:25

08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 34

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

High AC 0.300AC
~~0.300AC~~

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = CAVA 8164.14.00 09/16
05/25/2017 11:26

Test	AC	Time
01 Room Air	0.000	11:27
02 *Subject Test	INT*	11:27
03 Room Air	0.000	11:28

*Invalid Test
Interferent Detected

Sub Name = TEST, NA NA
Sub DOB = 12/12/1998
Sub Sex = Unknown Weight = NA
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Lot No = 01316080A2
Cyl No = 28
Expiration Date = 02/05/2018
County = 34 Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Interferent detected check 0.10AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = CAVA 8164.14.00 09/16
05/25/2017 11:28

Test	AC	Time
01 Diagnostic	OK	11:29
02 Room Air	RFI*	11:29
03 Room Air	0.000	11:30

*Invalid Test
Inhibited - RFI

Sub Name = RFI, TEST NA
Sub DOB = 01/01/2001
Sub Sex = Unknown Weight = NA
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Lot No = 01316080A2
Cyl No = 28
Expiration Date = 02/05/2018
County = 34 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = CAVA 8164.14.00 09/16
05/25/2017 11:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:11
02 Std. Gas	0.080	11:11
03 Room Air	0.000	11:12
04 Std. Gas	0.079	11:12
05 Room Air	0.000	11:13
06 Std. Gas	0.080	11:13
07 Room Air	0.000	11:14

Lot No = 01316080A2
Cyl No = 28
Exp Date = 02/05/2018
County = 34 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = CAVA 8164.14.00 09/16
05/25/2017 11:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:31
02 Std. Gas	0.079	11:31
03 Room Air	0.000	11:32
04 Std. Gas	0.080	11:32
05 Room Air	0.000	11:32
06 Std. Gas	0.080	11:33
07 Room Air	0.000	11:33

Lot No = 01316080A2
Cyl No = 28
Exp Date = 02/05/2018
County = 34 Oper No. = 777777

Deb Shanaver
Operator Signature
DEB SHANAVER

Remarks:
Calibration check 0.080 AC

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006491
Location = CAVA 8164.14.00 09/16
05/25/2017 11:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:34
02 Std. Gas	0.080	11:34
03 Room Air	0.000	11:35
04 Std. Gas	0.079	11:35
05 Room Air	0.000	11:36
06 Std. Gas	0.079	11:36
07 Room Air	0.000	11:37

Lot No = 01316080A2

Cyl No = 28

Exp Date = 02/05/2018

County = 34

Oper No. = 777777

Deb Shanaver

Operator Signature
DEB SHANAVER

Remarks:

Calibration check 0.080 AC

Form 106-I8000