

INTOXILYZER® 8000 INSPECTION

Serial No.: 80-006497

Location: GF Air Base

- Check When Done
- A. Is the breath tube heated? ✓
  - B. Is the diagnostic check complete? ✓
  - C. Is the time, date, and year correct (re-set if necessary)? ✓
  - D. Print test. (Attach test record.) ✓
  - E. Low AC. Use  $\leq 0.03$  AC in ACA mode. (Attach test record.)  
Sim SN DR5113 Lot No. 201507F Rep AC 0.020 ✓
  - F. High AC. Use  $\geq 0.25$  AC in ACA mode. (Attach test record.)  
Sim SN DR3378 Lot No. 116060 Rep AC 0.300 ✓
  - G. Interferent check. Use 0.05% acetone in ethanol solution in ABA mode.  
(Attach test record.)  
Sim SN DR3803 Lot No. 1CS5 Solution 0.10ACT 0.05% acetone ✓
  - H. RFI check. Run CMS mode. Key radio during CMS test.  
(Attach test record.) ✓
  - I. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.) 0.080 AC

Lot No. 19615080A5 Cyl No. 30 Exp. Date 09/05/2017

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.081</u>	Test 6 <u>0.082</u>	Test 9 <u>0.081</u>

Average 0.081

J. Tank pressure, Level 3 Func D Sub G P: 994 psi Reg 1800 psi

K. Remarks/Maintenance record: Annual inspection

24 May 17  
Date

30 MAY 2017  
Date

Deb Shanaver  
Field Inspector's Signature

[Signature]  
Reviewed by

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006497  
Location = GFAF                8164.14.00 09/16  
05/24/2017                      13:08

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|  
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            19615080A5  
Standard Cyl #?            30  
Standard Expiration?       09/05/2017  
Oper No?                    777777

Flow Cal. Date:            08/19/2015  
Slope                      657  
Intercept                  -642041

IR Calibration Date:       08/19/2015  
                                        3um                          9um


-----  
0th Coef(\*100):        -16255                          -17993  
1st Coef(\*100):        263339                        136695  
2nd Coef(\*100):        5198                            1666  
H2O adj(mg/l\*10k):    766                            522

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
DEB SHANAVER

Remarks:



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006497  
Location = GFAF      8164.14.00 09/16  
05/24/2017      13:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:25
02 Std. Sol.	0.018	13:26
03 Room Air	0.000	13:27
04 Std. Sol.	0.019	13:27
05 Room Air	0.000	13:28
06 Std. Sol.	0.019	13:29
07 Room Air	0.000	13:29

08 Sim Temp = 34.0°C

Simul Ser No = DR5113

Std Sol No = 201507F

County = 65

Oper No. = 777777

Deb Shanaver

Operator Signature  
DEB SHANAVER

Remarks:

Low AC 0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006497  
Location = GFAF      8164.14.00 09/16  
05/24/2017      13:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:31
02 Std. Sol.	0.298	13:32
03 Room Air	0.000	13:33
04 Std. Sol.	0.299	13:33
05 Room Air	0.000	13:34
06 Std. Sol.	0.300	13:35
07 Room Air	0.000	13:35

08 Sim Temp = 34.0°C

Simul Ser No = DR3378

Std Sol No = 16060

County = 65

Oper No. = 777777

*Deb Shanaver*

Operator Signature  
DEB SHANAVER

Remarks:

*High AC 0.300AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-006497  
Location = GFAF                    8164.14.00 09/16  
05/24/2017                            13:37

Test	AC	Time
01 Room Air	0.000	13:38
02 *Subject Test	INT*	13:38
03 Room Air	0.000	13:39

\*Invalid Test  
Interferent Detected

Sub Name = TEST, NA NA  
Sub DOB = 08/24/1970  
Sub Sex = Unknown                    Weight = NA  
Test = DUI                              Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 19615080A5  
Cyl No = 30  
Expiration Date = 09/05/2017  
County = 65                              Oper No. = 777777

Deb Shanaver  
Operator Signature  
DEB SHANAVER

Remarks:

*Interferent detected check 0.10AC +  
0.05% acetone*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006497  
Location = GFAF                8164.14.00 09/16  
05/24/2017                        13:22

Test	AC	Time
01 Diagnostic	OK	13:23
02 Room Air	RFI*	13:23
03 Room Air	0.000	13:23

\*Invalid Test  
Inhibited - RFI

Sub Name = RFI, NA NA  
Sub DOB = 08/28/1967  
Sub Sex = Unknown                Weight = NA  
Test = DUI                         Cit = NA  
Dr. Lic. = ND/NA  
Lot No = 19615080A5  
Cyl No = 30  
Expiration Date = 09/05/2017  
County = 65                        Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Shanaver

Operator Signature  
DEB SHANAVER

Remarks:

RFI check

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006497  
Location = GFAF      8164.14.00 09/16  
05/24/2017      13:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:11
02 Std. Gas	0.081	13:11
03 Room Air	0.000	13:12
04 Std. Gas	0.081	13:12
05 Room Air	0.000	13:13
06 Std. Gas	0.081	13:13
07 Room Air	0.000	13:14

Lot No = 19615080A5

Cyl No = 30

Exp Date = 09/05/2017

County = 65

Oper No. = 777777

*Deb Shanaver*

Operator Signature  
DEB SHANAVER

Remarks:

*Calibration check 0.080AC*

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006497  
Location = GFAF      8164.14.00 09/16  
05/24/2017      13:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:14
02 Std. Gas	0.081	13:15
03 Room Air	0.000	13:15
04 Std. Gas	0.081	13:16
05 Room Air	0.000	13:16
06 Std. Gas	0.082	13:17
07 Room Air	0.000	13:17

Lot No = 19615080A5

Cyl No = 30

Exp Date = 09/05/2017

County = 65

Oper No. = 777777

*Deb Shanaver*

Operator Signature  
DEB SHANAVER

Remarks:

*Calibration check 0.080AC*

Form 106-I8000



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006497  
Location = GFAP      8164.14.00 09/16  
05/24/2017      13:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:18
02 Std. Gas	0.081	13:19
03 Room Air	0.000	13:19
04 Std. Gas	0.081	13:20
05 Room Air	0.000	13:20
06 Std. Gas	0.081	13:21
07 Room Air	0.000	13:21

Lot No = 19615080A5  
Cyl No = 30  
Exp Date = 09/05/2017  
County = 65      Oper No. = 777777

Deb Shanaver  
Operator Signature  
DEB SHANAVER

Remarks:  
Calibration check 0.080 AC