

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006493

Location: KILLDEER PD
(KDPD)

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 998 psi Regulator: 1000 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.020 AC
Sim. Ser #: DR3375 Lot #: 17160 Exp. Date: 6-3-19
 - b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: DR7111 Lot #: 17350 Exp. Date: 10-11-19

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR7089 Lot #: JCS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 39717080A3 Cyl #: 15 Exp. Date: 2-5-20

Test 1 <u>0.081</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.080</u>	Test 5 <u>0.081</u>	Test 8 <u>0.081</u>
Test 3 <u>0.080</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Breath Analyst Signature

[Signature]
NA

Date

4.18.2019
NA

Reviewed by

Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006493
Location = KDPD 8164.14.00 09/16
04/18/2019 12:08

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&* () _ + ?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&* () _ + ?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer (Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34917080A3
Standard Cyl #? 15
Standard Expiration? 02/05/2020
Oper No? 666666

Flow Cal. Date: 05/23/2018
Slope 640
Intercept -624033

IR Calibration Date: 05/23/2018
 3um 9um

0th Coef(*100): -19857 -14695
1st Coef(*100): 263752 135087
2nd Coef(*100): 3253 1458
H2O adj (mg/l*10k): 643 483

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006493
Location = KDPD 8164.14.00 09/16
04/18/2019 12:15

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:16
02 Std. Sol.	0.020	12:17
03 Room Air	0.000	12:17
04 Std. Sol.	0.020	12:18
05 Room Air	0.000	12:18
06 Std. Sol.	0.021	12:19
07 Room Air	0.000	12:20

08 Sim Temp = 34.0°C

Simul Ser No = DR3375

Std Sol No = 17160

County = 13

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

LOW AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006493
Location = KDPD 8164.14.00 09/16
04/18/2019 12:21

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:22
02 Std. Sol.	0.297	12:23
03 Room Air	0.000	12:23
04 Std. Sol.	0.299	12:24
05 Room Air	0.000	12:25
06 Std. Sol.	0.300	12:25
07 Room Air	0.000	12:26

08 Sim Temp = 34.0°C

Simul Ser No = DR7111

Std Sol No = 17350

County = 13

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006493
Location = KDPD 8164.14.00 09/16
04/18/2019 12:11

Test	AC	Time
01 Room Air	0.000	12:12
02 *Subject Test	INT*	12:13
03 Room Air	0.000	12:13

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS821456
Lot No = 34917080A3
Cyl No = 15
Expiration Date = 02/05/2020
County = 13 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006493
Location = KDPD 8164.14.00 09/16
04/18/2019 12:28

Test	AC	Time
01 Diagnostic	OK	12:29
02 Room Air	0.000	12:30
03 *Subject Test	RFI*	12:30
04 Room Air	0.000	12:30

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 34917080A3
Cyl No = 15
Expiration Date = 02/05/2020
County = 13 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006493
Location = KDPD 8164.14.00 09/16
04/18/2019 12:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:05
02 Std. Gas	0.081	12:06
03 Room Air	0.000	12:06
04 Std. Gas	0.080	12:07
05 Room Air	0.000	12:07
06 Std. Gas	0.080	12:08
07 Room Air	0.000	12:08

Lot No = 34917080A3
Cyl No = 15
Exp Date = 02/05/2020
County = 13 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006493
Location = KDPD 8164.14.00 09/16
04/18/2019 12:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:32
02 Std. Gas	0.081	12:33
03 Room Air	0.000	12:33
04 Std. Gas	0.081	12:33
05 Room Air	0.000	12:34
06 Std. Gas	0.081	12:34
07 Room Air	0.000	12:35

Lot No = 34917080A3
Cyl No = 15
Exp Date = 02/05/2020
County = 13 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006493
Location = KDPD 8164.14.00 09/16
04/18/2019 12:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:36
02 Std. Gas	0.081	12:36
03 Room Air	0.000	12:37
04 Std. Gas	0.081	12:37
05 Room Air	0.000	12:38
06 Std. Gas	0.081	12:38
07 Room Air	0.000	12:38

Lot No = 34917080A3
Cyl No = 15
Exp Date = 02/05/2020
County = 13 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.0804K

Form 106-I8000