

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4952

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 491 psi Regulator: 500 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: DR3378 Lot #: 201712B Exp. Date: 12.12.19
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.380 AC
Sim. Ser #: DR7350 Lot #: 2018034 Exp. Date: 3-22-20

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 34418080A2 Cyl #: 45 Exp. Date: 2.5.21

Test 1	<u>0.079</u>	Test 4	<u>0.078</u>	Test 7	<u>0.079</u>
Test 2	<u>0.078</u>	Test 5	<u>0.078</u>	Test 8	<u>0.079</u>
Test 3	<u>0.079</u>	Test 6	<u>0.079</u>	Test 9	<u>0.079</u>
Average	<u>0.079</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION
CALIBRATED TANK SENSOR

Instrument is acceptable to be used in the field.

Charles E. Eden
Breath Analyst Signature

6/18/2019
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/18/2019 08:42

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 34418080A2
Standard Cyl #? 45
Standard Expiration? 02/05/2021
Oper No? 666666

Flow Cal. Date: 01/19/2010
Slope 694
Intercept -638616

IR Calibration Date: 05/18/2016
 3um 9um

0th Coef(*100): -17808 -20871
1st Coef(*100): 256729 129583
2nd Coef(*100): 2826 1310
H2O adj(mg/l*10k): 344 422

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/18/2019 08:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:47
02 Std. Sol.	0.019	08:48
03 Room Air	0.000	08:48
04 Std. Sol.	0.019	08:49
05 Room Air	0.000	08:49
06 Std. Sol.	0.018	08:50
07 Room Air	0.000	08:50

08 Sim Temp = 34.0°C

Simul Ser No = DR3378
Std Sol No = 201712B
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Low AC
0.020 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/18/2019 09:02

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:02
02 Std. Sol.	0.297	09:03
03 Room Air	0.000	09:04
04 Std. Sol.	0.296	09:04
05 Room Air	0.000	09:05
06 Std. Sol.	0.296	09:06
07 Room Air	0.000	09:06

08 Sim Temp = 34.0°C

Simul Ser No = DR7350
Std Sol No = 201803H
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.300 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/18/2019 09:10

Test	AC	Time
01 Room Air	0.000	09:11
02 *Subject Test	INT*	09:12
03 Room Air	0.000	09:12

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 45
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/18/2019 10:20

Test	AC	Time
01 Diagnostic	OK	10:21
02 Room Air	0.000	10:21
03 *Subject Test	RFI*	10:21
04 Room Air	0.000	10:22

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 34418080A2
Cyl No = 45
Expiration Date = 02/05/2021
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks: RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/18/2019 10:23

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:23
02 Std. Gas	0.079	10:24
03 Room Air	0.000	10:24
04 Std. Gas	0.078	10:24
05 Room Air	0.000	10:25
06 Std. Gas	0.079	10:25
07 Room Air	0.000	10:26

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/18/2019 10:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:28
02 Std. Gas	0.078	10:28
03 Room Air	0.000	10:28
04 Std. Gas	0.078	10:29
05 Room Air	0.000	10:29
06 Std. Gas	0.079	10:30
07 Room Air	0.000	10:30

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = TOXL 8164.14.00 09/16
06/18/2019 10:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:31
02 Std. Gas	0.079	10:31
03 Room Air	0.000	10:32
04 Std. Gas	0.079	10:32
05 Room Air	0.000	10:33
06 Std. Gas	0.079	10:33
07 Room Air	0.000	10:34

Lot No = 34418080A2
Cyl No = 45
Exp Date = 02/05/2021
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000