

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-006685

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 855 psi Regulator: 850 psi

B. Tests (Sign and attach test records):

- Configure simulator for the following tests (Level 1,S).
1. Wet Calibration Check
 - a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.015 AC
Sim. Ser #: DR5113 Lot #: 201805C Exp. Date: 5.30.20
 - b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.250 AC
Sim. Ser #: DR7351 Lot #: 201803G Exp. Date: 3.22.20

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.100 AC
Gas Cyl Lot #: 13518100 43 Cyl #: 6 Exp. Date: 8.5.20

Test 1	<u>0.101</u>	Test 4	<u>0.101</u>	Test 7	<u>0.102</u>
Test 2	<u>0.101</u>	Test 5	<u>0.102</u>	Test 8	<u>0.102</u>
Test 3	<u>0.101</u>	Test 6	<u>0.102</u>	Test 9	<u>0.102</u>
Average	<u>0.102</u>				

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Charles E. Ee
Breath Analyst Signature

6.26.2019
Date

NA
Reviewed by

NA
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006685
Location = TOXL 8164.14.00 09/16
06/26/2019 15:38

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.100
Standard Lot #? 13518100A3
Standard Cyl #? 6
Standard Expiration? 08/05/2020
Oper No? 666666

Flow Cal. Date: 09/06/2016
Slope 705
Intercept -687370

IR Calibration Date: 09/06/2016
 3um 9um

	3um	9um
0th Coef(*100):	-14155	-22915
1st Coef(*100):	271436	135296
2nd Coef(*100):	2016	1251
H2O adj(mg/l*10k):	372	505

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks: *PRINT TEST*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006685
Location = TOXL 8164.14.00 09/16
06/26/2019 15:40

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:41
02 Std. Sol.	0.013	15:41
03 Room Air	0.000	15:42
04 Std. Sol.	0.013	15:43
05 Room Air	0.000	15:43
06 Std. Sol.	0.013	15:44
07 Room Air	0.000	15:44

08 Sim Temp = 34.0°C

Simul Ser No = DR5113

Std Sol No = 201805C

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

Low AC
0.015 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006685
Location = TOXL 8164.14.00 09/16
06/26/2019 15:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:53
02 Std. Sol.	0.246	15:54
03 Room Air	0.000	15:55
04 Std. Sol.	0.245	15:55
05 Room Air	0.000	15:56
06 Std. Sol.	0.245	15:57
07 Room Air	0.000	15:57


08 Sim Temp = 34.0°C

Simul Ser No = DR7351

Std Sol No = 201803G

County = 08

Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

HIGH AC
0.250 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006685
Location = TOXL 8164.14.00 09/16
06/26/2019 15:59

Test	AC	Time
01 Room Air	0.000	16:00
02 *Subject Test	INT*	16:00
03 Room Air	0.000	16:01

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 13518100A3
Cyl No = 6
Expiration Date = 08/05/2020
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006685
Location = TOXL 8164.14.00 09/16
06/26/2019 16:02

Test	AC	Time
01 Diagnostic	OK	16:03
02 Room Air	0.000	16:03
03 *Subject Test	RFI*	16:03
04 Room Air	0.000	16:04

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 13518100A3
Cyl No = 6
Expiration Date = 08/05/2020
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006685
Location = TOXL 8164.14.00 09/16
06/26/2019 16:04

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:05
02 Std. Gas	0.101	16:05
03 Room Air	0.000	16:06
04 Std. Gas	0.101	16:06
05 Room Air	0.000	16:07
06 Std. Gas	0.101	16:07
07 Room Air	0.000	16:08

Lot No = 13518100A3
Cyl No = 6
Exp Date = 08/05/2020
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CALIBRATION CHECK*
0.100 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006685
Location = TOXL 8164.14.00 09/16
06/26/2019 16:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:09
02 Std. Gas	0.101	16:09
03 Room Air	0.000	16:10
04 Std. Gas	0.102	16:10
05 Room Air	0.000	16:11
06 Std. Gas	0.102	16:11
07 Room Air	0.000	16:12

Lot No = 13518100A3
Cyl No = 6
Exp Date = 08/05/2020
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: *CEE*
AT CALIBRATION CHECK
0.100 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006685
Location = TOXL 8164.14.00 09/16
06/26/2019 16:12

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:13
02 Std. Gas	0.102	16:13
03 Room Air	0.000	16:14
04 Std. Gas	0.102	16:14
05 Room Air	0.000	16:15
06 Std. Gas	0.102	16:15
07 Room Air	0.000	16:16

Lot No = 13518100A3
Cyl No = 6
Exp Date = 08/05/2020
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.100 AC

Form 106-I8000