



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Chris Pulver</u>		
Location: <u>New Town PD</u>		
Intoxilyzer® Serial Number: <u>80-004951</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>36</u>	Expiration Date: <u>12-23-11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>079</u> % Test 2: 0. <u>078</u> % Test 3: 0. <u>078</u> % <div style="text-align: right;">(ATTACH TEST RECORD)</div>		

Chris Pulver
Chemical Test Operator Signature

1-9-11
Date

Deb Kashner
Reviewed By

12 Jan 2011
Date


Intoxilyzer Test Record and Checklist
NDCAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004951
Location = NEWT 8164.13.00 06/09
01/09/2011 00:45

DRY CAL CHECK

Test	PC	Time
01 Room Air	0.000	00:47
02 Std. Gas	0.079	00:47
03 Room Air	0.000	00:48
04 Std. Gas	0.078	00:48
05 Room Air	0.000	00:49
06 Std. Gas	0.078	00:49
07 Room Air	0.000	00:49

Lot No = 709352
Cyl No = 36
Exp Date = 12/23/2011
County = 31 Oper No. = 049912


Operator Signature
CHRISTOPHER PULVER

Remarks:

Form 106-18000

80-004951
P. 56 of 56
JPM