

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Mark W. McDonald

Location: Valley City Police Dept.

Intoxilyzer Serial No.: 40-00 4184

Lot Number: 659358

Cylinder Number: 47

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.029 %

Test 2: 0.081 %

Test 3: 0.081 %

(ATTACH TEST RECORD)

[Signature]  
Chemical Test Operator Signature

2/15/10  
Date

Deb Keshur  
Reviewed By

10 Mar 2010  
Date

80-004184  
20/21 FM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004184  
Location = VALL      8164.13.00 06/09  
02/05/2010      10:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:54
02 Std. Gas	0.079	10:55
03 Room Air	0.000	10:55
04 Std. Gas	0.081	10:56
05 Room Air	0.000	10:56
06 Std. Gas	0.081	10:57
07 Room Air	0.000	10:57

Lot No = 659358  
Cyl No = 47  
Exp Date = 09/03/2011  
County = 02      Oper No. = 119502



Operator Signature  
MARK W. MCDONALD

Remarks:

Form 106-I8000

80-004184  
7/21 KM