



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Brenda Janssen				
Location:	Kenville County Sheriff Office				
Intoxilyzer® Serial Number:	80-003069				
Gas Lot Number:	659358	Gas Cylinder Number:	27	Expiration Date:	9/3/11
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: 0. <u>080</u> %					
Test 2: 0. <u>081</u> %					
Test 3: 0. <u>081</u> %					
(ATTACH TEST RECORD)					

Brenda Janssen
Chemical Test Operator Signature

Feb 28, 2011
Date

Deb Kashner
Reviewed By

01 Mar 2011
Date

80-003069
P. 9/1/92
JFM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = MOHL 8164.13.00 06/09
02/28/2011 11:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:26
02 Std. Gas	0.080	11:26
03 Room Air	0.000	11:27
04 Std. Gas	0.081	11:27
05 Room Air	0.000	11:27
06 Std. Gas	0.081	11:28
07 Room Air	0.000	11:28

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 011102


Operator Signature
BRENDA K. JANSSEN

Remarks:

Form 106-I8000

80-003069
P. 92 of 92
SM