

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David J. Zess

Location: Cassher County Sheriff's Office (Langdon)

Intoxilyzer Serial No.: 80-004188

Lot Number: 659358

Cylinder Number: 53

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.079 %

Test 2: 0.079 %

Test 3: 0.079 %

(ATTACH TEST RECORD)

David J. Zess
Chemical Test Operator Signature

03/19/10
Date

Deb Kashner
Reviewed By

23 Mar 2010
Date

80-004188
P. 1 of 2

26/77


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004188
Location = LANG 8164.13.00 06/09
03/19/2010 16:01

DRY CAL CHECK

| Test | AC | Time |
|-------------|-------|-------|
| 01 Room Air | 0.000 | 16:02 |
| 02 Std. Gas | 0.079 | 16:02 |
| 03 Room Air | 0.000 | 16:03 |
| 04 Std. Gas | 0.079 | 16:03 |
| 05 Room Air | 0.000 | 16:04 |
| 06 Std. Gas | 0.079 | 16:04 |
| 07 Room Air | 0.000 | 16:05 |

Lot No = 659358
Cyl No = 53
Exp Date = 09/03/2011
County = 10 Oper No. = 067518



Operator Signature
DAVID J. ZEIS

Remarks:

Form 106-18000

80-004188
p. 2 of 2
7/7/11