



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Sterling Reed		
Location:	New Town		
Intoxilyzer® Serial Number:	80-004951		
Gas Lot Number:	709352	Gas Cylinder Number:	36
		Expiration Date:	12-23-11
<input checked="" type="checkbox"/> New Cylinder Installed <input type="checkbox"/> Cylinder Not Changed			
Instrument Test Results (Report to 3 Digits; Example, 0.081%)			
Test 1: 0.079 %			
Test 2: 0.079 %			
Test 3: 0.078 %			
(ATTACH TEST RECORD)			

Sterling Reed
Chemical Test Operator Signature

3-19-10
Date

Deb Kashur
Reviewed By

29 Mar 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004951
Location = NEWT 8164.13.00 06/09
03/29/2010 21:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:01
02 Std. Gas	0.079	21:02
03 Room Air	0.000	21:02
04 Std. Gas	0.079	21:02
05 Room Air	0.000	21:03
06 Std. Gas	0.078	21:03
07 Room Air	0.000	21:04

Lot No = 709352
Cyl No = 36
Exp Date = 12/23/2011
County = 60 Oper No. = 020717


Operator Signature
STERLING R. REED

Remarks:

80-004951
33/33 JRM