



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Lawson Johnny Z</i>		
Location <i>Bowman Co SO</i>	Intoxilyzer® Serial Number <i>80-004188</i>	
Gas Lot Number <i>16512080A2</i>	Gas Cylinder Number <i>31</i>	Gas Expiration Date <i>08/01/2014</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0. 078 AC
 - B. 0. 077 AC
 - C. 0. 078 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Johnny Z Lawson</i>	Date <i>05-29-2014</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kasher</i>	Date <i>03 June 2013</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004188
Location = BOWM 8164.13.00 06/09
05/29/2013 08:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:48
02 Std. Gas	0.078	08:48
03 Room Air	0.000	08:49
04 Std. Gas	0.077	08:49
05 Room Air	0.000	08:50
06 Std. Gas	0.078	08:50
07 Room Air	0.000	08:51

Lot No = 16512080A2
Cyl No = 31
Exp Date = 08/01/2014
County = 06 Oper No. = 020814



Operator Signature
JOHNNY LAWSON

Remarks: