



ETHANOL BREATH STANDARD CYLINDER REPORT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name:	Margy Pearson				
Location:	Ransom Co SD, Lisbon, ND				
Intoxilyzer® Serial Number:	80-004952				
Gas Lot Number:	709352	Gas Cylinder Number:	22	Expiration Date:	12/13/2011
<input checked="" type="checkbox"/> New Cylinder Installed	<input type="checkbox"/> Cylinder Not Changed				
Instrument Test Results (Report to 3 Digits; Example, 0.081%)					
Test 1: <u>0.078</u> %					
Test 2: <u>0.078</u> %					
Test 3: <u>0.079</u> %					
(ATTACH TEST RECORD)					

Margy Pearson
Chemical Test Operator Signature

2 June 2010
Date

Deb Kashur
Reviewed By

14 June 2010
Date

80-004952
P. 34 of 35
JAN


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004952
Location = LISB 8164.13.00 06/09
06/02/2010 17:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:48
02 Std. Gas	0.078	17:48
03 Room Air	0.000	17:49
04 Std. Gas	0.078	17:49
05 Room Air	0.000	17:50
06 Std. Gas	0.079	17:50
07 Room Air	0.000	17:50

Lot No = 709352
Cyl No = 22
Exp Date = 12/23/2011
County = 37 Oper No. = 107501



Operator Signature
MARGY PEARSON

Remarks:

ACA Cyl Install

Form 106-I8000

80-004952
P. 35 of 35
JRM