



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <i>Wingers, Matthew</i>		
Location <i>Bldg # 547 - Minot Air Force Base <sup>DK</sup></i>	Intoxilyzer® Serial Number <i>80-004196</i>	
Gas Lot Number <i>13314080A2</i>	Gas Cylinder Number <i>55</i>	Gas Expiration Date <i>6/5/16</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. 0.078 AC  
 B. 0.079 AC  
 C. 0.079 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature <i>Matthew Wingers</i>	Date <i>8/4/14</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kashner</i>	Date <i>13 Aug. 2014</i>

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004196  
Location = MNAF                8164.13.00 06/09  
08/04/2014                        17:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:46
02 Std. Gas	0.078	17:47
03 Room Air	0.000	17:47
04 Std. Gas	0.079	17:48
05 Room Air	0.000	17:48
06 Std. Gas	0.079	17:49
07 Room Air	0.000	17:49

Lot No = 13314080A2  
Cyl No = 53  
Exp Date = 06/05/2016  
County = 64                      Oper No. = 130777

  
\_\_\_\_\_  
Operator Signature  
MATTHEW WINGERS

Remarks: