



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Berry Vannatta</u>		
Location: <u>Moh L</u>		
Intoxilyzer® Serial Number: <u>80-603069</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>27</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>079</u> %		
Test 2: 0. <u>079</u> %		
Test 3: 0. <u>086</u> %		
(ATTACH TEST RECORD)		

Berry Vannatta  
Chemical Test Operator Signature

10-21-10  
Date

Deb Kashum  
Reviewed By

11 Jan 2011  
Date

80-003069  
P. 83 of 84  
FM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003069  
Location = MOHL      8164.13.00 06/09  
10/21/2010      12:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:46
02 Std. Gas	0.079	12:46
03 Room Air	0.000	12:47
04 Std. Gas	0.079	12:47
05 Room Air	0.000	12:48
06 Std. Gas	0.080	12:48
07 Room Air	0.000	12:49

Lot No = 659358  
Cyl No = 27  
Exp Date = 09/03/2011  
County = 38      Oper No. = 039909

  
Operator Signature  
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-003069  
P. 84 of 84  
KM