



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>DUISEN, SHANNON P.</u>		
Location: <u>TIOGA</u>		
Intoxilyzer® Serial Number: <u>80-003071</u>		
Gas Lot Number: <u>25410080A1</u>	Gas Cylinder Number: <u>25</u>	Expiration Date: <u>7/01/12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.078</u> %		
Test 2: <u>0077</u> %		
Test 3: <u>0078</u> %		
(ATTACH TEST RECORD)		

[Signature]
Chemical Test Operator Signature

10/24/10
Date

Deb Kashur
Reviewed By

27 Oct 2010
Date

80-003071
P. 100 of 101
FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003071
Location = TIOG 8164.13.00 06/09
10/24/2010 01:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:11
02 Std. Gas	0.078	01:11
03 Room Air	0.000	01:11
04 Std. Gas	0.077	01:12
05 Room Air	0.000	01:12
06 Std. Gas	0.078	01:13
07 Room Air	0.000	01:13

Lot No = 15410080A1
Cyl No = 25
Exp Date = 07/01/2012
County = 53 Oper No. = 020632



Operator Signature
SHANNON P. DUISEN

Remarks:

ACA

Form 106-I8000

80-003071
P. 101 of 101
FM