



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Barry Vaughan</u>		
Location: <u>Mohl</u>		
Intoxilyzer® Serial Number: <u>80-003069</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>27</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>080</u> %		
Test 3: 0. <u>679</u> %		
(ATTACH TEST RECORD)		

Barry Vaughan  
 Chemical Test Operator Signature

11-08-10  
 Date

Deb Kashur  
 Reviewed By

16 NOV 2010  
 Date

80-003069  
 P. 77 of 78  
 YKH

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003069  
Location = MOHL      8164.13.00 06/09  
11/08/2010      16:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:07
02 Std. Gas	0.080	16:07
03 Room Air	0.000	16:08
04 Std. Gas	0.080	16:08
05 Room Air	0.000	16:08
06 Std. Gas	0.079	16:09
07 Room Air	0.000	16:09


Lot No = 659358

Cyl No = 27

Exp Date = 09/03/2011

County = 38

Oper No. = 039909



Operator Signature  
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-003069  
P. 78 of 78  
JCM