



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Dustin Plummer</u>		
Location: <u>MohL</u>		
Intoxilyzer® Serial Number: <u>80-003069</u>		
Gas Lot Number: <u>059358</u>	Gas Cylinder Number: <u>27</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>080</u> %		
Test 2: 0. <u>081</u> %		
Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Dustin Plummer
Chemical Test Operator Signature

12-06-10
Date

Deb Kashner
Reviewed By

10 Dec '10
Date

80-003069
P. 79 of 80
JAN


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = MOHL 8164.13.00 06/09
12/06/2010 12:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:54
02 Std. Gas	0.080	12:55
03 Room Air	0.000	12:55
04 Std. Gas	0.081	12:55
05 Room Air	0.000	12:56
06 Std. Gas	0.080	12:56
07 Room Air	0.000	12:57

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 100914




Operator Signature
DUSTIN J. PLUMMER

Remarks:

Form 106-I8000

I, the undersigned Sheriff in and for Benville
County, North Dakota, do certify that the
foregoing is a true copy of an original document
remaining on file and re-held in my office.

WITNESS my hand and seal of said Court this
06 day of December, 2010.
Dustin Plummer deputy, Sheriff


80-003069
P. 80 of 80
JAN