

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Richard D. Johnson

Location: McLean S.O. Washburn

Intoxilyzer Serial No.: 80-004196

Lot Number: 659357

Cylinder Number: 47

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.081 %

Test 2: 0.081 %

Test 3: 0.082 %

(ATTACH TEST RECORD)

Richard D. Johnson  
Chemical Test Operator Signature

12/29/09  
Date

Deb Kasher (ACA-check)  
Reviewed By

05 Jan 2010  
Date

80-004196  
P. 1 of 2  
JRN

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004196  
Location = WASH      8164.13.00 06/09  
12/29/2009      21:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:48
02 Std. Gas	0.081	21:49
03 Room Air	0.000	21:49
04 Std. Gas	0.081	21:50
05 Room Air	0.000	21:50
06 Std. Gas	0.082	21:50
07 Room Air	0.000	21:51


Lot No = 659357

Cyl No = 47

Exp Date = 09/03/2011

County = 28

Oper No. = 118707



Operator Signature  
RICHARD D. JOHNSON

Remarks:

Form 106-I8000

80-004196  
P. 2 of 2 JRM