



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) HUARD, CHRISTOPHER J		
Location RZCHLAND COUNTY JAZZ	Intoxilyzer® Serial Number 80-006493	
Gas Lot Number 02621080A1	Gas Cylinder Number 033	Gas Expiration Date 03/05/2023

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 A. **0.080** AC
 B. **0.081** AC
 C. **0.080** AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Christopher Huard</i>	132570	Date 06/11/21
Reviewed By (Crime Laboratory Use Only) <i>Charles E. Ehr</i>		Date 6.17.21

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006493
Location = WAHP 8164.14.00 09/16
06/11/2021 21:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:21
02 Std. Gas	0.080	21:21
03 Room Air	0.000	21:22
04 Std. Gas	0.081	21:22
05 Room Air	0.000	21:23
06 Std. Gas	0.080	21:23
07 Room Air	0.000	21:24

Lot No = 02621080A1
Cyl No = 33
Exp Date = 03/05/2023
County = 39 Oper No. = 132570


Operator Signature
CHRISTOPHER HUARD

Remarks: **INSTALL**

Form 106-I8000