



**ETHANOL GAS STANDARD CYLINDER REPORT**  
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM  
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Meadows, Cedy D.</i>		
Location <i>Garrison</i>	Intoxilyzer® Serial Number <i>80-006685</i>	
Gas Lot Number <i>26021080A1</i>	Gas Cylinder Number <i>3d</i>	Gas Expiration Date <i>10/25/2023</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).  
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).  
 A. *0.081* AC  
 B. *0.081* AC  
 C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:  
 A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).  
 B. ACA Test Record.

*05*  
*CEE*  
*11-4-21*

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>11/3/21</i>
Reviewed By (Crime Laboratory Use Only) <i>[Signature]</i>	Date <i>11-4-2021</i>


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-006685  
Location = GARR                    8164.14.00 09/16  
11/03/2021                            16:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:19
02 Std. Gas	0.081	16:19
03 Room Air	0.000	16:20
04 Std. Gas	0.081	16:20
05 Room Air	0.000	16:21
06 Std. Gas	0.082	16:21
07 Room Air	0.000	16:22

Lot No = 26021080A1  
Cyl No = 32  
Exp Date = 10/05/2023  
County = 28                            Oper No. = 100724



Operator Signature  
CODY MEADOWS

Remarks:

Form 106-I8000