

CHAIN OF CUSTODY  
BREATH ALCOHOL INSTRUMENT

Instrument: Intox 8000 Serial Number: 80-006493

Agency being assigned instrument: CM1

Date 22 Jan 16

Transferred from:

Transferred to:

Name (print) Deb Shanaver

Name (print) CM1

Signature Deb Shanaver

Signature \_\_\_\_\_

Agency Crime Lab

Agency \_\_\_\_\_

Reason for transfer/remarks:

Pressing F9 (<sup>fan</sup>~~flow~~) does not turn on the fan during  
initial inspection. Under Warranty.

Or Shipped Via:

USPS \_\_\_\_\_ UPS \_\_\_\_\_ FedEx \_\_\_\_\_ Other \_\_\_\_\_

Date Sending Instrument: \_\_\_\_\_ / \_\_\_\_\_ / \_\_\_\_\_ Air Bill No. \_\_\_\_\_