

BREATH ALCOHOL INSTRUMENT CHAIN OF CUSTODY WORKSHEET

Instrument: 8000 Serial Number: 80-004951

Transported Out of the Crime Laboratory:

Agency Being Assigned Instrument: _____

Name/Agency of Person Transporting Instrument:

_____ (Print Name) _____ (Print Agency)

Signature of Person Receiving Instrument: _____

Date Receiving Instrument: ____/____/____ From: _____

Or Shipped Via:

USPS ____ UPS ____ FedEx ____ Other _____

Date Sending Instrument: ____/____/____ Air Bill No. _____

Complete Air Bill (Identify Instrument/Serial Number) and attach a copy to this worksheet.

Transported ^{to} the Crime Laboratory:

Agency Returning Instrument: CM

Name/Agency of Person Returning Instrument:

_____ (Print Name) _____ (Print Agency)

Signature of Person Returning Instrument: _____

Date Returning Instrument: 12/13 July 2011 Received By: Deb Keshur

Or Received Via:

USPS ____ UPS ____ FedEx ____ Other _____

Date Receiving Instrument: ____/____/____ Air Bill No. _____

Attach shipping paperwork (if any) to this worksheet.

Instrument Transfer Information Entered Into the Equipment Database:

Date of Action 12 July 2011 By: Deb Keshur

80-004951
P. 80 of 82
km



SERVICE WORK ORDER

DATE	5-17-11
RCVD VIA	FEDX

316 E. 9th St., Owensboro KY 42303
Phone 866-835-0690
Fax: 270-685-6268

326642

BILL TO:		
ATTN:		
CITY	STATE	ZIP
CUSTOMER NO.	585013	
PHONE		

SHIP TO: ND Crime Lab		
ATTN: Del Kashur		
2641 E. Main Ave.		
CITY	STATE	ZIP
Bismarck	ND	58501
PHONE	701-328-6159	
FAX	701-328-6185	

MODEL #	8000	SERIAL #	80-004951	BILL CODE	ELC
EXTRA PARTS RCVD BT, gas box, battery					
DESCRIPTION OF PROBLEM EEPROM FAILURE					

Infrared / Fuel Cell PRELIMINARY TEST INFORMATION	BATTERY- EXT 2.99V; INT 3.0V
VERIFIED PROBLEM? EEPROM FAILURE YES	
WORK PERFORMED: Replaced EEPROM, installed 110211 cap.	
SEAL # 04044	
Miscellaneous PRELIMINARY TEST INFORMATION	
WORK PERFORMED: 	

PARTS USED		
QTY	PART NO.	DESCRIPTION
1	220286	EEPROM
1	110211	cap.

REPAIR	CAL. ADJ.	CALIBRATION	FINAL
KM/H	DR	DR	DR
1/2 HOURS	HOURS	HOURS	HOURS
16	1	.5	.5
DATE	DATE	DATE	DATE
6/23/11	6/27	7/1	7/5

SHIPPING STICKER HERE
SPECIAL SHIPPING INSTRUCTIONS
80-004951 P. 81 of 82 PM

Certificate of Calibration

Intoxilyzer® 8000

*This is to certify that the calibration of **Intoxilyzer**® serial number 80-004951, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.*

Date

July 1, 2011

Signed

[Signature]

Technician

CMI INC.

80-004951

P. 82482
PM

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650519 CMI 7/9/09