

BREATH ALCOHOL INSTRUMENT CHAIN OF CUSTODY WORKSHEET

Instrument: 8000 Serial Number: 80-4188 ⁸⁰⁻⁰⁰⁴¹⁸⁸ ^{OK}

Transported Out of the Crime Laboratory:

Agency Being Assigned Instrument: _____

Name/Agency of Person Transporting Instrument:

_____ (Print Name) _____ (Print Agency)

Signature of Person Receiving Instrument: _____

Date Receiving Instrument: ____/____/____ From: _____

Or Shipped Via:

USPS ____ UPS ____ FedEx ____ Other _____

Date Sending Instrument: ____/____/____ Air Bill No. _____

Complete Air Bill (Identify Instrument/Serial Number) and attach a copy to this worksheet.

Transported to the Crime Laboratory:

Agency Returning Instrument: CMU

Name/Agency of Person Returning Instrument:

_____ (Print Name) _____ (Print Agency)

Signature of Person Returning Instrument: _____

Date Returning Instrument: 19 July 2011 Received By: Ashley Ayd

Or Received Via:

USPS ____ UPS ____ FedEx ____ Other _____

Date Receiving Instrument: ____/____/____ Air Bill No. _____

Attach shipping paperwork (if any) to this worksheet.

Instrument Transfer Information Entered Into the Equipment Database:

Date of Action 19 July 2011 By: Deb Kashner



SERVICE WORK ORDER

DATE	5-17-11
RCVD VIA	FEDX

316 E. 9th St., Owensboro KY 42303
 Phone 866-835-0690
 Fax: 270-685-6268

326643

BILL TO:
ATTN:
CITY STATE ZIP
CUSTOMER NO. 585013
PHONE

SHIP TO: ND Crime Lab
ATTN: Deb Kashur
2641 E. Main Ave.
CITY STATE ZIP
Bismarck ND 58501
PHONE 701-328-6159
FAX 701-328-6185

MODEL # 8000	SERIAL # 80-004188	BILL CODE EW
EXTRA PARTS RCVD BF, 8000 box, battery - BU-1168		
DESCRIPTION OF PROBLEM		

Infrared / Fuel Cell PRELIMINARY TEST INFORMATION	BATTERY-EXT <u>DISC V</u> ; INT <u>2.04 V</u>
VERIFIED PROBLEM? <u>EEPROM FAIL</u>	
WORK PERFORMED: <u>Replaced EEPROM.</u> <u>Installed Cap PER CM11-076</u> <u>Replaced external battery.</u>	
SEAL # <u>04047</u>	
Miscellaneous PRELIMINARY TEST INFORMATION	
WORK PERFORMED: <u>BA</u>	

PARTS USED		
QTY	PART NO.	DESCRIPTION
1	220286	EEPROM
1	110211	Cap.
1	013010	Battery 12V 38A

REPAIR	CAL. ADJ.	CALIBRATION	FINAL
LH.	DR	DR	DR
1/2 HOURS	HOURS	HOURS	HOURS
07	1	5	5
DATE	DATE	DATE	DATE
7-5-11	7-7-11	7-7-11	7-7-11

SHIPPING STICKER HERE
SPECIAL SHIPPING INSTRUCTIONS
80-004188
P. 133 of 134
PM

Certificate of Calibration

Intoxilyzer® 8000

*This is to certify that the calibration of **Intoxilyzer**® serial number 80-004188, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.*

Date July 7, 2011

Signed 
Technician

CMI INC.

80-004188
P. 134 of 134
JRM

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650519 CMI 7/9/09