

CHAIN OF CUSTODY
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-003069

Agency being assigned instrument: Crime Lab

Date 25 Sept 2013

Transferred from:

Transferred to:

Name (print) CMI

Name (print) Deb Kashur

Signature _____

Signature Deb Kashur

Agency _____

Agency Crime Lab

Reason for transfer/remarks:

CMI Replaced Intox

Or Shipped Via:

USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: _____ / _____ / _____ Air Bill No. _____



SERVICE WORK ORDER

316 E. 9th St., Owensboro KY 42303
 Phone 866-835-0690
 Fax: 270-685-6268

330576

DATE: 11-27-12
 RCVD VIA: Fed Ex

BILL TO:

ATTN:

CITY STATE ZIP

CUSTOMER NO.

PHONE

SHIP TO: NO State of

ATTN: Cumi Lab

2641 E Main Ave

CITY STATE ZIP
 Bismark ND 58501

PHONE: 701-328-6159

FAX: 701-328-6185

MODEL # 8000 SERIAL # 80-003069 BILL CODE 1

EXTRA PARTS RCVD: Gas Thief

DESCRIPTION OF PROBLEM: Hard to blow through

Infrared / Fuel Cell
 PRELIMINARY TEST INFORMATION

BATTERY-EXT V; INT 3.03 V

Flow Sensor low R value
 Battery not charging
 Voltage current failure

VERIFIED PROBLEM? Yes

WORK PERFORMED:
 Replaced Flow Sensor for low R value.
 Replaced bad battery.
 Replaced Power Distribution board for Voltage current failures.

PARTS USED		
QTY	PART NO.	DESCRIPTION
1	210120	Flow Sensor
1	021510	Battery Pack
1	021288	Power Dist Bd.

Miscellaneous
 PRELIMINARY TEST INFORMATION

WORK PERFORMED:
 VOT

REPAIR	CAL. ADJ.	CALIBRATION	FINAL
JMT	DP	DP	DR
1/2 HOURS	HOURS	HOURS	HOURS
10	1	.5	.5
DATE	DATE	DATE	DATE
Sept 19, 2013	Sept 17, 2013	Sept 17, 2013	Sept 18, 2013

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS