

CHAIN OF CUSTODY  
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-007092

Agency being assigned instrument: UND P.D.

Date 5.14.19

Transferred from:

Transferred to:

Name (print) CHARLES E. EDER

Name (print) Tracy L. Meidinger

Signature Charles Eder

Signature Tracy L. Meidinger

Agency O.A.G. - CRIME LAB

Agency UND-Police Dept.

Reason for transfer/remarks:

- REPLACING 80-005358
- 80-005358 READING 0.016 AC ON 0.020 AC STANDARD. STILL WITHIN +/- 0.005 TOLERANCE

Or Shipped Via:

USPS \_\_\_\_\_ UPS \_\_\_\_\_ FedEx \_\_\_\_\_ Other \_\_\_\_\_

Date Sending Instrument: \_\_\_\_\_ / \_\_\_\_\_ / \_\_\_\_\_ Air Bill No. \_\_\_\_\_