



CHAIN OF CUSTODY  
BREATH ALCOHOL INSTRUMENT

Instrument: 8000 Serial Number: 80-006493

Agency being assigned instrument: O.A.G.-CRIME LAB- BREATH

Date 2/8/22

<u>Transferred from:</u>	<u>Transferred to:</u>
Name (print) <u>CHRISTOPHER HUARD</u>	Name (print) <u>CHARLES E EDER</u>
Signature <u></u>	Signature <u></u>
Agency <u>WAHPETON P.D.</u>	Agency <u>O.A.G.-CRIME LAB-BREATH</u>

Reason for transfer/remarks:

- REPORTED INSTRUMENT IS NOT RESPONDING TO ANY INPUT. SCREEN<sup>OFF</sup> LOCKED UP SATURDAY (2/5/22).
- REBOOTING DOES NOT HELP. CANNOT CONNECT REMOTELY.
- BEING REPLACED BY 80-006508

Or Shipped Via:

USPS \_\_\_\_\_ UPS \_\_\_\_\_ FedEx \_\_\_\_\_ Other \_\_\_\_\_

Date Sending Instrument: \_\_\_\_\_ / \_\_\_\_\_ / \_\_\_\_\_ Air Bill No. \_\_\_\_\_