

CHAIN OF CUSTODY  
BREATH ALCOHOL INSTRUMENT

Instrument: Intoxilyzer 8000 Serial Number: 80-004952

Agency being assigned instrument: Griggs CO SO

Date 5/17/2022

Transferred from:

Transferred to:

Name (print) Amberly Muehler

Name (print) Janelle Pertschiller

Signature [Signature]

Signature [Signature]

Agency Griggs

Agency NDAG-CLD

Reason for transfer/remarks:

*Transported to CLD for Annual Inspection.  
EEPROM error message upon arrival at CLD.  
will replace with ~~80-004936~~ <sup>JP</sup> 80-004956*

Or Shipped Via:

USPS  UPS  FedEx  Other

Date Sending Instrument:      /      /      Air Bill No.