



SERVICE WORK ORDER

316 E. 9th St., Owensboro KY 42303
Phone 866-835-0690
Fax: 270-685-6268

338125

DATE	1-25-16
RCVD VIA	FEDX

BILL TO:		
ATTN:		
CITY	STATE	ZIP
CUSTOMER NO.	585013	
PHONE		

SHIP TO: ND Crime Laboratory		
ATTN:		
2641 E Main St.		
CITY	STATE	ZIP
Kismock	ND	58501
PHONE	701-328-6159	
FAX	701-328-6185	

MODEL #	8000	SERIAL #	80-006493	BILL CODE	3
---------	------	----------	-----------	-----------	---

EXTRA PARTS RCVD
PC, gas box, Battery, BT

DESCRIPTION OF PROBLEM
Fan doesn't work when F9 is pressed

Infrared / Fuel Cell
PRELIMINARY TEST INFORMATION

BATTERY- EXT 13.76v; INT 3.02v

Fan does not turn on

VERIFIED PROBLEM?
yes

WORK PERFORMED:

Reseated Fan plug
Performed system and update checks
GTW

Miscellaneous
PRELIMINARY TEST INFORMATION

WORK PERFORMED:

100

PARTS USED		
QTY	PART NO.	DESCRIPTION

REPAIR	CAL. ADJ.	CALIBRATION	FINAL
WEP	WEP	WEP	WEP
½ HOURS	HOURS	HOURS	HOURS
3	0.5	1.0	1.0
DATE	DATE	DATE	DATE
Feb 1, 2016	Feb 12, 16	Feb 12, 16	Feb 12, 16

SHIPPING STICKER HERE

SPECIAL SHIPPING INSTRUCTIONS