



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59281 (03-2014)

Serial Number <u>80-004951</u>	Instrument Location <u>NEWT (BIA FT. BERTHOLD T.A.T.)</u>
Reason for Install/Repair <input type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>INSTALL DURING ANNUAL INSPECTION</u>	

Check When Done:

- ☒ 1. Surge Protector Installed/Properly Grounded.
- ☒ NA 2. Telephone Line Connected to Intoxilyzer® 8000. NO ANALOG LINE
- ☒ 3. Breath Tube Heated.
- ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 6. Run Tests:
 - A. Print Test (Level 1, Function P). ☒
 - B. ACA Test (Level 1, Function C). ☒
 - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test). ☒
- ☒ NA 7. Repair and/or Maintenance Performed (if Any): INSTALL DURING ANNUAL INSPECTION
- ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <u>Charles E. Eder</u>	Date <u>21 MAY 15</u>
Reviewed By (Crime Laboratory Use Only) <u>Deb Keshen</u>	Date <u>21 May 15</u>

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: Charles E. Eder Certified Date: 29 Nov. 2016

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004951
Location = NEWT 8164.13.00 06/09
05/21/2015 12:48

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 33913080A4
Standard Cyl #? 23
Standard Expiration? 01/01/2016
Oper No? 666666

Flow Cal. Date: 06/27/2011
Slope 704
Intercept -660586

IR Calibration Date: 05/20/2015
 3um 9um

0th Coef(*100):	-12971	-20926
1st Coef(*100):	253059	131517
2nd Coef(*100):	2934	1439
H2O adj(mg/l*10k):	464	531

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINT TEST

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004951
Location = NEWT 8164.13.00 06/09
05/21/2015 12:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:50
02 Std. Gas	0.080	12:50
03 Room Air	0.000	12:51
04 Std. Gas	0.080	12:51
05 Room Air	0.000	12:52
06 Std. Gas	0.079	12:52
07 Room Air	0.000	12:53

Lot No = 33913080A4
Cyl No = 23
Exp Date = 01/01/2016
County = 60 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

ACA TEST

0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004951
Location = NEWT 8164.13.00 06/09
05/21/2015 12:54

Test	AC	Time
01 Diagnostic	OK	12:56
02 Room Air	0.000	12:56
03 *Subject Test	RFI*	12:56
04 Room Air	0.000	12:57

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS821456
Lot No = 33913080A4
Cyl No = 23
Expiration Date = 01/01/2016
County = 60 Oper No. = 666666

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK