



# INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59281 (03-2014)

Serial Number <u>80-006495</u>	Instrument Location <u>Fargo Police Dept PD3</u>
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

### Check When Done:

- 1. Surge Protector Installed/Properly Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
  - A. Print Test (Level 1, Function P) ✓
  - B. ACA Test (Level 1, Function C) ✓
  - C. RFI Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if Any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer®8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <u>[Signature]</u>	Date <u>8-2-16 ✓</u>
Reviewed By (Crime Laboratory Use Only) <u>Deb Sheavee</u>	Date <u>08 Aug 16</u>

This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Certified By: [Signature]      Certified Date: 01 Nov. 16

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer          Alcohol Analyzer  
North Dakota Model 8000          SN 80-006495  
Location = FAPD                    8164.13.00 06/09  
08/02/2016                            14:51

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstu vwxyz1234567890- = |  
ABCDEFGHIJKLMN O PQRSTU VWXYZ!@# \$%^&\*() \_ + ?

abcdefghijklmnopqrstu vwxyz1234567890- = |  
ABCDEFGHIJKLMN O PQRSTU VWXYZ!@# \$%^&\*() \_ + ?

Current Instrument Setup


Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                Yes  
Inhib Printer(Y/N)?                No  
Display Volume?                    No  
Disable On Memfull?                Yes  
# of Print Copies?                    1  
Select Std (D/W/I)?                Dry  
Standard Value?                    0.080  
Standard Lot #?                    09216080A3  
Standard Cyl #?                    37  
Standard Expiration?                05/05/2018  
Oper No?                            110104

Flow Cal. Date:                    08/18/2015  
Slope                                678  
Intercept                            -657244

IR Calibration Date:                08/18/2015

	3um	9um
0th Coef(*100):	-22405	-18664
1st Coef(*100):	270469	136654
2nd Coef(*100):	3040	1360
H2O adj(mg/l*10k):	699	443

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
\_\_\_\_\_  
Operator Signature  
TROY NIELSEN

Remarks: *Print test*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006495  
Location = FAPD      8164.13.00 06/09  
08/02/2016      14:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:52
02 Std. Gas	0.080	14:53
03 Room Air	0.000	14:53
04 Std. Gas	0.080	14:54
05 Room Air	0.000	14:54
06 Std. Gas	0.080	14:55
07 Room Air	0.000	14:55

Lot No = 09216080A3  
Cyl No = 37  
Exp Date = 05/05/2018  
County = 09      Oper No. = 110104

  
\_\_\_\_\_  
Operator Signature  
TROY NIELSEN

Remarks:      A/A Test

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006495  
Location = FAPD      8164.13.00 06/09  
08/02/2016      14:56

Test	AC	Time
01 Room Air	RFI*	14:57
02 Room Air	0.000	14:57

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, TEST N  
Sub DOB = 01/01/1990  
Sub Sex = Male      Weight = 250  
Test = DUI      Cit = 123456  
Dr. Lic. = ND/TES908765  
Lot No = 09216080A3  
Cyl No = 37  
Expiration Date = 05/05/2018  
County = 09      Oper No. = 110104

  
\_\_\_\_\_  
Operator Signature  
TROY NIELSEN

Remarks:      RFI test