











































Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003069  
Location = TOXL      8164.14.00 09/16  
11/16/2018      12:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:00
02 Std. Gas	0.081	12:01
03 Room Air	0.000	12:01
04 Std. Gas	0.080	12:02
05 Room Air	0.000	12:02
06 Std. Gas	0.081	12:02
07 Room Air	0.000	12:03

Lot No = 03417080A3  
Cyl No = 7  
Exp Date = 03/05/2019  
County = 08      Oper No. = 888888

Roberta Grieger-Nimmo  
Operator Signature  
ROBERTA GRIEGER-NIMMO

Remarks: Dry Calibration Check  
0.080 AC

Form 106-I8000