



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-006490	Instrument Location Tioga PD (27 2nd St Ne)
Reason for Install/Repair	
<input type="checkbox"/> Install After Receiving From Crime Laboratory <input checked="" type="checkbox"/> Install After Location Change <input checked="" type="checkbox"/> Other (Specify) <u>Install into new Police Department Building</u>	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): None
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>John Nelson</i>	Date 04/14/2020
---	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <i>Chanda E. Edr</i>	Certified Date 4/16/20
---	----------------------------------

Note: Booking area is in the process of obtaining an analog data line. Step 2 is incomplete at this time until data line is established L

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TIOG 8164.14.00 09/16
04/14/2020 16:36

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 24119080A1
Standard Cyl #? 2
Standard Expiration? 11/05/2021
Oper No? 130977

Flow Cal. Date: 08/18/2015
Slope 675
Intercept -635799

IR Calibration Date: 08/18/2015
 3um 9um

0th Coef(*100): -18073 -15729
1st Coef(*100): 265337 133887
2nd Coef(*100): 3965 1560
H2O adj(mg/l*10k): 767 457

***** Printer Test End *****



Operator Signature
JOSHUA NELSON

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TIOG 8164.14.00 09/16
04/14/2020 16:38

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:39
02 Std. Gas	0.083	16:39
03 Room Air	0.000	16:40
04 Std. Gas	0.083	16:40
05 Room Air	0.000	16:41
06 Std. Gas	0.082	16:41
07 Room Air	0.000	16:42

Lot No = 24119080A1
Cyl No = 2
Exp Date = 11/05/2021
County = 53 Oper No. = 130977


Operator Signature
JOSHUA NELSON

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TIOG 8164.14.00 09/16
04/14/2020 16:45

Test	AC	Time
01 Room Air	RFI*	16:46
02 Room Air	0.000	16:47

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST TEST
Sub DOB = 01/01/1988
Sub Sex = Male Weight = 125
Test = DUI Cit = 00000000
Dr. Lic. = ND/TEST
Lot No = 24119080A1
Cyl No = 2
Expiration Date = 11/05/2021
County = 53 Oper No. = 130977



Operator Signature
JOSHUA NELSON

Remarks:

Form 106-I8000