

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006493
Location = RUGB 8164.14.00 09/16
06/15/2020 16:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	16:39
02 Room Air	0.000	16:40

*Invalid Test
Inhibited - RFI

Lot No = 13518080A6
Cyl No = 23
Exp Date = 08/05/2020
County = 35 Oper No. = 130392



Operator Signature
DARREN HEIDBREDER

Remarks:

Form 106-18000