

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006496
Location = CASS 8164.13.00 06/09
06/18/2020 15:06

| Test | AC | Time |
|------------------|-------|-------|
| 01 Room Air | 0.000 | 15:08 |
| 02 *Subject Test | RFI* | 15:08 |
| 03 Room Air | 0.000 | 15:09 |

*Invalid Test
Inhibited - RFI

Sub Name = RFI TEST, RFI TEST RFI TEST
Sub DOB = 06/18/2000
Sub Sex = Male Weight = 1
Test = DUI Cit = RFI TEST
Dr. Lic. = ND/RFI TEST
Lot No = 17618080A3
Cyl No = 73
Expiration Date = 08/05/2020
County = 09 Oper No. = 050910



Operator Signature
CHAD THOMPSON

Remarks:

RFI Test

Form 106-I8000