

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006495
Location = FAPD 8164.14.00 09/16
06/22/2020 12:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:56
02 Std. Gas	0.081	12:57
03 Room Air	RFI*	12:57
04 Room Air	0.000	12:57

*Invalid Test
Inhibited - RFI

Lot No = 34418080A2
Cyl No = 10
Exp Date = 02/05/2021
County = 09 Oper No. = 110104


Operator Signature
TROY NIELSEN

Remarks:

Form 106-I8000