







Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-006491  
Location = LAMO      8164.14.00 09/16  
07/02/2020      18:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:22
02 Std. Gas	0.077	18:23
03 Room Air	RFI*	18:23
04 Room Air	0.000	18:24

\*Invalid Test  
Inhibited - RFI

Lot No = 05620080A1  
Cyl No = 12  
Exp Date = 04/05/2022  
County = 23      Oper No. = 132539



Operator Signature  
JOSHUA FASTNACHT

Remarks:

Form 106-I8000