

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-007098

Location: TOXL

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 358 psi Regulator: 375 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value \leq 0.03 AC: 0.015 AC
Sim. Ser #: MP3064 Lot #: 201805C Exp. Date: 5-30-20
- b. High AC (Level 1,C): Known Value \geq 0.25 AC: 0.300 AC
Sim. Ser #: MP3069 Lot #: 201902D Exp. Date: 2-20-21

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR3451 Lot #: ICS 5 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 05620080A | Cyl #: 47 Exp. Date: 4-5-22
Test 1 0.081 Test 4 0.079 Test 7 0.081
Test 2 0.081 Test 5 0.082 Test 8 0.082
Test 3 0.082 Test 6 0.080 Test 9 0.081
Average 0.081

C. Remarks/Maintenance: ANNUAL INSPECTION

Instrument is acceptable to be used in the field.

Breath Analyst Signature

Charles E. Ed

NA

Date

5/5/2020

NA

Reviewed by

Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
05/05/2020 11:36

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* ()_+?

abcdefghijklmnopqrstuvwxy1234567890- = |
ABCDEFGHIJKLMN OPQRSTUVWXYZ!@# \$%^&* ()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 05620080A1
Standard Cyl #? 47
Standard Expiration? 04/05/2022
Oper No? 666666

Flow Cal. Date: 09/26/2018
Slope 684
Intercept -588446

IR Calibration Date: 09/26/2018
 3um 9um

0th Coef(*100): -40948 -15551
1st Coef(*100): 237555 134140
2nd Coef(*100): 2571 1433
H2O adj(mg/l*10k): 1097 457

***** Printer Test End *****



Operator Signature
CHARLES EDER

Remarks:

PRINTTEST

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
05/05/2020 11:38

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:38
02 Std. Sol.	0.013	11:39
03 Room Air	0.000	11:39
04 Std. Sol.	0.013	11:40
05 Room Air	0.000	11:41
06 Std. Sol.	0.012	11:41
07 Room Air	0.000	11:42

08 Sim Temp = 34.0°C

Simul Ser No = MP3064
Std Sol No = 201805C
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

LOW AC
0.015 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
05/05/2020 12:06

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:06
02 Std. Sol.	0.293	12:07
03 Room Air	0.000	12:08
04 Std. Sol.	0.294	12:08
05 Room Air	0.000	12:09
06 Std. Sol.	0.295	12:10
07 Room Air	0.000	12:10


08 Sim Temp = 34.0°C

Simul Ser No = MP3069

Std Sol No = 201902B

County = 08

Oper No. = 666666



Operator Signature

CHARLES EDER

Remarks:

HIGH AC

0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
05/05/2020 13:05

Test	AC	Time
01 Room Air	0.000	13:06
02 *Subject Test	INT*	13:06
03 Room Air	0.000	13:07

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

Interferent Check
0.10 AC + 0.05% ACETONE

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
05/05/2020 13:03

Test	AC	Time
01 Diagnostic	OK	13:03
02 Room Air	0.000	13:04
03 *Subject Test	RFI*	13:04
04 Room Air	0.000	13:04

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CHECK
Dr. Lic. = ND/DIS921456
Lot No = 05620080A1
Cyl No = 47
Expiration Date = 04/05/2022
County = 08 Oper No. = 666666

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
CHARLES EDER

Remarks:

RFI CHECK

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
05/05/2020 13:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:07
02 Std. Gas	0.081	13:08
03 Room Air	0.000	13:08
04 Std. Gas	0.081	13:09
05 Room Air	0.000	13:09
06 Std. Gas	0.082	13:09
07 Room Air	0.000	13:10

Lot No = 05620080A1
Cyl No = 47
Exp Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
05/05/2020 13:11

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:11
02 Std. Gas	0.079	13:11
03 Room Air	0.000	13:12
04 Std. Gas	0.082	13:12
05 Room Air	0.000	13:13
06 Std. Gas	0.080	13:13
07 Room Air	0.000	13:14

Lot No = 05620080A1
Cyl No = 47
Exp Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks:

CALIBRATION CHECK
0.080 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
05/05/2020 13:14

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:15
02 Std. Gas	0.081	13:15
03 Room Air	0.000	13:15
04 Std. Gas	0.082	13:16
05 Room Air	0.000	13:16
06 Std. Gas	0.081	13:17
07 Room Air	0.000	13:17

Lot No = 05620080A1
Cyl No = 47
Exp Date = 04/05/2022
County = 08 Oper No. = 666666



Operator Signature
CHARLES EDER

Remarks: CALIBRATION CHECK
0.080 AC

Form 106-I8000