

INTOXILYZER® 8000 INSPECTION

Intoxilyzer® 8000 Serial Number: 80-007698

Location: Tux L
80-007098

A. General Setup and Checks:

1. Diagnostic Tests Pass and Instrument in Ready Mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
 - a. Display: 952 psi Regulator: 950 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a. Low AC (Level 1,C): Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP3070 Lot #: 22CP41184 Exp. Date: 3/1/24
- b. High AC (Level 1,C): Known Value ≥ 0.25 AC: 0.400 AC
Sim. Ser #: MP3065 Lot #: 202103E Exp. Date: 3/24/23

Configure dry gas standard for the following tests (Level 1,S).

2. Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone
Sim. Ser #: DR5114 Lot #: IC56 Exp. Date: NA
 Display reads "Interferent Detect"
3. RFI Check (CMS Mode) Display reads "RFI Detect"
4. Dry Calibration Check (Level 1,C): Known Value 0.080 AC
Gas Cyl Lot #: 02621080A1 Cyl #: 2 Exp. Date: 3/5/2023

Test 1 <u>0.080</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.080</u>	Test 8 <u>0.080</u>
Test 3 <u>0.081</u>	Test 6 <u>0.080</u>	Test 9 <u>0.080</u>
Average <u>0.081</u>		

C. Remarks/Maintenance: Annual Inspection

Replaced breath tube hose o-ring. Replaced battery.

Instrument is acceptable to be used in the field.

Janelle Portschtler
Breath Analyst Signature

04/04/2022
Date

[Signature]
Reviewed by

04/22
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
04/04/2022 15:27

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 02621080A1
Standard Cyl #? 19
Standard Expiration? 03/05/2023
Oper No? 132128

Flow Cal. Date: 09/26/2018
Slope 684
Intercept -588446

IR Calibration Date: 09/26/2018
 3um 9um

0th Coef(*100): -40948 -15551
1st Coef(*100): 237555 134140
2nd Coef(*100): 2571 1433
H2O adj(mg/l*10k): 1097 457

***** Printer Test End *****



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Print test*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
04/04/2022 15:46

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:46
02 Std. Sol.	0.018	15:47
03 Room Air	0.000	15:48
04 Std. Sol.	0.019	15:48
05 Room Air	0.000	15:49
06 Std. Sol.	0.020	15:50
07 Room Air	0.000	15:50

08 Sim Temp = 34.0°C

Simul Ser No = MP3070
Std Sol No = 22CP41184
County = 08 Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Low Ac 0.020*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
04/04/2022 15:51

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:52
02 Std. Sol.	0.393	15:53
03 Room Air	0.000	15:53
04 Std. Sol.	0.397	15:54
05 Room Air	0.000	15:54
06 Std. Sol.	0.397	15:55
07 Room Air	0.000	15:56

08 Sim Temp = 34.0°C

Simul Ser No = MP3065

Std Sol No = 202103E

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks:

High AC ~~0.000~~ 0.400


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
04/04/2022 15:57

Test	AC	Time
01 Room Air	0.000	15:58
02 *Subject Test	INT*	15:58
03 Room Air	0.000	15:59

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = INTERFERENT CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128



Operator Signature
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
04/04/2022 15:59

Test	AC	Time
01 Diagnostic	OK	16:00
02 Room Air	0.000	16:00
03 *Subject Test	RFI*	16:01
04 Room Air	0.000	16:01

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 01/01/1982
Sub Sex = Female Weight = 150
Test = DUI Cit = RFI CK
Dr. Lic. = ND/DIS821456
Lot No = 02621080A1
Cyl No = 2
Expiration Date = 03/05/2023
County = 08 Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
JANELLE PORTSCHELLER

Remarks: *RFI check*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
04/04/2022 16:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:02
02 Std. Gas	0.080	16:03
03 Room Air	0.000	16:03
04 Std. Gas	0.081	16:04
05 Room Air	0.000	16:04
06 Std. Gas	0.081	16:04
07 Room Air	0.000	16:05

Lot No = 02621080A1
Cyl No = 2
Exp Date = 03/05/2023
County = 08 Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
04/04/2022 16:05

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:06
02 Std. Gas	0.081	16:06
03 Room Air	0.000	16:07
04 Std. Gas	0.080	16:07
05 Room Air	0.000	16:07
06 Std. Gas	0.080	16:08
07 Room Air	0.000	16:08

Lot No = 02621080A1

Cyl No = 2

Exp Date = 03/05/2023

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 4-6*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TOXL 8164.16.00 09/18
04/04/2022 16:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:09
02 Std. Gas	0.081	16:09
03 Room Air	0.000	16:10
04 Std. Gas	0.080	16:10
05 Room Air	0.000	16:11
06 Std. Gas	0.080	16:11
07 Room Air	0.000	16:12

Lot No = 02621080A1

Cyl No = 2

Exp Date = 03/05/2023

County = 08

Oper No. = 132128

Janelle Portscheller

Operator Signature
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check # 7-9*

Form 106-I8000