



ETHANOL GAS STANDARD CYLINDER REPORT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Josh Nelson</i>		
Location <i>Tigoga PD</i>	Intoxilyzer® Serial Number <i>80-007098</i>	
Gas Lot Number <i>02621080A1</i>	Gas Cylinder Number <i>19</i>	Gas Expiration Date <i>03/05/2023</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
- 12 Apr 22* Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.080* AC
 - B. *0.081* AC
 - C. *0.082* AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Josh Nelson</i>	Date <i>04/06/2022</i>
Reviewed By (Crime Laboratory Use Only) <i>Janelle Putschell</i>	Date <i>12 Apr 22</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007098
Location = TIOG 8164.16.00 09/18
04/06/2022 17:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:45
02 Std. Gas	0.080	17:45
03 Room Air	0.000	17:45
04 Std. Gas	0.081	17:46
05 Room Air	0.000	17:46
06 Std. Gas	0.082	17:47
07 Room Air	0.000	17:47

Lot No = 02621080A1
Cyl No = 19
Exp Date = 03/05/2023
County = 53 Oper No. = 130977



Operator Signature
JOSHUA NELSON

Remarks:

Form 106-I8000