

**INTOXILYZER® 8000 INSPECTION**

Intoxilyzer® 8000 Serial Number: 80-0070917 Location: TOXL

A. General Setup and Checks:

1.  Diagnostic Tests Pass and Instrument in Ready Mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G). Display and Regulator within 50 psi.
  - a. Display: 586 psi Regulator: 600 psi

B. Tests (Sign and attach test records):

Configure simulator for the following tests (Level 1,S).

1. Wet Calibration Check

- a.  Low AC (Level 1,C): Known Value  $\leq 0.03$  AC: 0.020 AC  
Sim. Ser #: MP 3060 Lot #: 202201A Exp. Date: 1/11/24
- b.  High AC (Level 1,C): Known Value  $\geq 0.25$  AC: 0.300 AC  
Sim. Ser #: MP 3064 Lot #: 202012A Exp. Date: 12/8/22

Configure dry gas standard for the following tests (Level 1,S).

2.  Interferent Check (Level 1,B): Known Value: 0.10 AC + 0.05% Acetone  
Sim. Ser #: DR3847 Lot #: ICSC Exp. Date: NA  
 Display reads "Interferent Detect"
3.  RFI Check (CMS Mode)  Display reads "RFI Detect"
4.  Dry Calibration Check (Level 1,C): Known Value 0.080 AC  
Gas Cyl Lot #: 02621080A1 Cyl #: 4 Exp. Date: 3/5/2023

Test 1	<u>0.082</u>	Test 4	<u>0.081</u>	Test 7	<u>0.082</u>
Test 2	<u>0.082</u>	Test 5	<u>0.081</u>	Test 8	<u>0.081</u>
Test 3	<u>0.081</u>	Test 6	<u>0.081</u>	Test 9	<u>0.082</u>
Average	<u>0.081</u>				

C. Remarks/Maintenance: Annual Inspection

Instrument is acceptable to be used in the field.

Janelle Portschiller  
Breath Analyst Signature

4/5/2022  
Date

[Signature]  
Reviewed by

041222  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-007097  
Location = TOXL                    8164.16.00 09/18  
04/05/2022                            10:20

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPNOPQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:                    Enabled  
Start Test Sequence:                DABACABA  
Display Prelim Rslt?                Yes  
Display Third Digit?                Yes  
Inhib Printer(Y/N)?                No  
Display Volume?                    No  
Disable On Memfull?                Yes  
# of Print Copies?                  1  
Select Std (D/W/I)?                Dry  
Standard Value?                    0.080  
Standard Lot #?                    05620080A1  
Standard Cyl #?                    10  
Standard Expiration?                04/05/2022  
Oper No?                            132128

Flow Cal. Date:                    09/03/2020  
Slope                                660  
Intercept                            -636264

IR Calibration Date:                09/03/2020  
    3um                                9um

-----  
0th Coef(\*100):            -36810                            -25919  
1st Coef(\*100):            254044                            139068  
2nd Coef(\*100):            1858                                1258  
H2O adj (mg/l\*10k):        903                                584

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
JANELLE PORTSCHELLER

Remarks:                    *Print test*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-007097  
Location = TOXL                    8164.16.00 09/18  
04/05/2022                            10:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:26
02 Std. Sol.	0.020	10:28
03 Room Air	0.000	10:28
04 Std. Sol.	0.020	10:29
05 Room Air	0.000	10:30
06 Std. Sol.	0.021	10:30
07 Room Air	0.000	10:31

08 Sim Temp = 34.0°C

Simul Ser No = MP3060  
Std Sol No = 202201A  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks:            *Low AC 0.020*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007097  
Location = TOXL      8164.16.00 09/18  
04/05/2022      10:33

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:33
02 Std. Sol.	0.304	10:34
03 Room Air	0.000	10:34
04 Std. Sol.	0.308	10:35
05 Room Air	0.000	10:36
06 Std. Sol.	0.313	10:36
07 Room Air	0.000	10:37

08 Sim Temp = 34.0°C

Simul Ser No = MP3064

Std Sol No = 202012A

County = 08

Oper No. = 132128

*Janelle Portscheller*

Operator Signature  
JANELLE PORTSCHELLER

Remarks:

*High AC 0.300*

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-007097  
Location = TOXL                    8164.16.00 09/18  
04/05/2022                            10:45

Test	AC	Time
01 Room Air	0.000	10:46
02 *Subject Test	INT*	10:47
03 Room Air	0.000	10:47

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = INTERFERENT CK  
Dr. Lic. = ND/DIS821456  
Lot No = 02621080A1  
Cyl No = 4  
Expiration Date = 03/05/2023  
County = 08                            Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Interferent Check*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-007097  
Location = TOXL                8164.16.00 09/18  
04/05/2022                            10:49

Test	AC	Time
01 Diagnostic	OK	10:50
02 Room Air	0.000	10:51
03 *Subject Test	RFI*	10:51
04 Room Air	0.000	10:52

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 01/01/1982  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = RFI CK  
Dr. Lic. = ND/DIS821456  
Lot No = 02621080A1  
Cyl No = 4  
Expiration Date = 03/05/2023  
County = 08                            Oper No. = 132128

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *RFI Check*


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007097  
Location = TOXL      8164.16.00 09/18  
04/05/2022      10:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:52
02 Std. Gas	0.082	10:53
03 Room Air	0.000	10:53
04 Std. Gas	0.082	10:54
05 Room Air	0.000	10:54
06 Std. Gas	0.081	10:54
07 Room Air	0.000	10:55

Lot No = 02621080A1  
Cyl No = 4  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128

  
\_\_\_\_\_  
Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #1-3*





Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-007097  
Location = TOXL      8164.16.00 09/18  
04/05/2022      11:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:01
02 Std. Gas	0.082	11:01
03 Room Air	0.000	11:01
04 Std. Gas	0.081	11:02
05 Room Air	0.000	11:02
06 Std. Gas	0.082	11:03
07 Room Air	0.000	11:03

Lot No = 02621080A1  
Cyl No = 4  
Exp Date = 03/05/2023  
County = 08      Oper No. = 132128



Operator Signature  
JANELLE PORTSCHELLER

Remarks: *Dry Calibration Check #7-9*